Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST LUCIE COUNTY SO
Time of Inspection: 13:10

Date of Inspection: 11/15/2017

Serial Number: 80-000788
Software: 8100.27

BLC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.082	0.198	0.081
0.000	0.050	0.082	0.200	0.081
0.000	0.050	0.082	0.201	0.081
0.000	0.050	0.082	0.200	0.082
0.000	0.049	0.083	0.200	0.082
0.000	0.050	0.082	0.200	0.082
0.000	0.049	0.083	0.200	0.082
0.000	0.050	0.082	0.200	0.082
0.000	0.050	0.083	0.200	0.082
0.000	0.050	0.082	0.200	0.082
Standard Deviations	0.0004	0.0004	0.0007	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

11/15/2017
Date

11/15/17
SD

RECEIVED
SEP 07 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST LUCIE COUNTY SO
Time of Inspection: 11:50
Date of Inspection: 09/07/2017
Serial Number: 80-000788
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.082	0.201	0.081
0.000	0.049	0.082	0.202	0.081
0.000	0.050	0.082	0.202	0.081
0.000	0.050	0.082	0.202	0.082
0.000	0.050	0.082	0.201	0.081
0.000	0.050	0.082	0.202	0.082
0.000	0.050	0.082	0.202	0.082
0.000	0.050	0.082	0.202	0.082
0.000	0.050	0.082	0.202	0.082
0.000	0.050	0.082	0.202	0.081
Standard Deviations	0.0004	0.0000	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt
SHAYLA D PLATT
Signature and Printed Name

09/07/2017
Date

9/7/17

RECEIVED
MAY 16 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: St. Lucie County SO
Time of Inspection:
Date of Inspection: 05/09/2017
Serial Number: 80-000788
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	yes		Mouth Alcohol Test: Slope Not Met	yes	
Interferent Detect Test: Interferent Detect	yes		Diagnostic Check (Post-Inspection): OK SP	~~yes~~	No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
Standard Deviations				

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks: Inadvertently un-plugged instrument during dry gas standard testing. Prior to losing power the dry gas standard test produced a 0.094 and INT. ALL OTHER VALUES WERE WITHIN RANGE. SP

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Plath SHAYLA PLATH
Signature and Printed Name

5/9/17
Date

RECEIVED
MAY 16 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST LUCIE COUNTY SO
Time of Inspection: 13:56 Date of Inspection: 05/16/2017

Serial Number: 80-000788 BCC
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.079	0.196	0.082
0.000	0.051	0.080	0.200	0.081
0.000	0.051	0.080	0.200	0.081
0.000	0.051	0.079	0.200	0.081
0.000	0.051	0.080	0.199	0.081
0.000	0.052	0.080	0.199	0.081
0.000	0.052	0.079	0.199	0.081
0.000	0.051	0.080	0.199	0.081
0.000	0.051	0.080	0.199	0.081
0.000	0.052	0.080	0.199	0.082

Standard Deviations	0.0008	0.0004	0.0011	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt SHAYLA D PLATT
Signature and Printed Name

05/16/2017
Date