

RECEIVED

JUL 27 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 14:01

Date of Inspection: 07/03/2017

Serial Number: 80-000944
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.081	0.198	0.078
0.000	0.051	0.081	0.199	0.078
0.000	0.051	0.081	0.199	0.079
0.000	0.051	0.081	0.199	0.078
0.000	0.051	0.081	0.200	0.078
0.000	0.051	0.081	0.200	0.078
0.000	0.051	0.081	0.200	0.077
0.000	0.051	0.081	0.200	0.077
0.000	0.051	0.081	0.199	0.078

Standard Deviations	0.0004	0.0003	0.0009	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes-Rivera

Signature and Printed Name DAVID E REYES-RIVERA

07/03/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

JUL 27 2017
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 09:27

Date of Inspection: 07/03/2017

Serial Number: 80-000944
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

✓ BUC

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.051	0.081	0.200	0.079
0.000	0.051	0.081	0.202	0.078
0.000	0.051	0.080	0.202	0.078
0.000	0.052	0.079	0.202	0.077
0.000	0.051	0.079	0.202	0.078
0.000	0.051	0.077	0.203	0.078
0.000	0.051	0.077	0.202	0.077
0.000	0.052	0.076	0.202	0.077
0.000	0.051	0.076	0.202	0.077
0.000	0.052	0.075	0.202	0.077

Standard Deviations	0.0004	0.0021	0.0007	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0009 Number of Simulators Used: 5

Pgm

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes-Rivera

Signature and Printed Name

DAVID E REYES-RIVERA

DR

07/03/2017
Date

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

MAR 28 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 12:43

Date of Inspection: 03/16/2017

Serial Number: 80-000944
Software: 8100.27

BK

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.051	0.082	0.201	0.082
0.000	0.051	0.082	0.202	0.081
0.000	0.051	0.082	0.203	0.081
0.000	0.051	0.083	0.203	0.080
0.000	0.052	0.083	0.203	0.081
0.000	0.051	0.083	0.203	0.080
0.000	0.051	0.083	0.203	0.080
0.000	0.051	0.082	0.204	0.080
0.000	0.051	0.083	0.204	0.080
0.000	0.052	0.082	0.203	0.080

Standard Deviations	0.0004	0.0005	0.0008	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

DFM

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

03/16/2017
Date