Florida Department of Law Enforcement **Alcohol Testing Program**

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Liberty County S.O. Serial Number: 80-000960 Time of Inspection: Date of Inspection: 11/8/2017 Software: 8100.27 Check or Test YES NO Check or Test YES NO Diagnostic Check Date and/or Time Adjusted No (Pre-Inspection): OK No Minimum Sample Volume Barometric Pressure Sensor No No Check: OK Check: OK Alcohol Free Subject Mouth Alcohol Test: No Test: 0.000 No Slope Not Met Interferent Detect Test: Diagnostic Check No No Interferent Detect (Post-Inspection): OK Alcohol Free 0.05g/210L Test 0.08g/210L Test 0.20g/210L Test 0.08 g/210L Test (g/210L)(g/210L)(g/210L) Dry Gas Std Test (g/210L) Lot#: Lot#: Lot#: (g/210L)Exp: Exp: Exp: Lot#: Exp: Standard Deviations Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: Number of Simulators Used: Instrument has power supply problem preventing the administration of the full Department Inspection. Instrument is being sent to CMI for repair. The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

> 11/8/2017 Date

Signature and Printed Name

Patrick J. Murphy

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BIC