

Florida Department of Law Enforcement RECEIVED

Alcohol Testing Program

AUG 04 2017

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: HAINES CITY PD
Time of Inspection: 17:19

Date of Inspection: 08/03/2017

Serial Number: 80-001007
Software: 8100.27

vcc

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.048	0.078 / 0.077	0.204	0.082
0.000	0.049	0.077 / 0.079	0.202	0.082
0.000	0.049	0.077 / 0.080	0.202	0.081
0.000	0.049	0.076 / 0.080	0.201	0.081
0.000	0.049	0.075 / 0.080	0.201	0.081
0.000	0.050	0.074 / 0.080	0.201	0.081
0.000	0.049	0.073 / 0.080	0.201	0.081
0.000	0.050	0.072 / 0.080	0.200	0.081
0.000	0.049	0.072 / 0.080	0.201	0.081
0.000	0.050	0.070 / 0.080	0.200	0.081
Standard Deviations	0.0006	0.0026 / 0.0009	0.0011	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance LOST POWER TO SIMULATOR.

Turned Sim back on, allowed to reheat then retested. DMS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE M BELL

Signature and Printed Name

JD
8/4/17

08/03/2017
Date