

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 02 2017

FDLE

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Alcohol Testing Program
BK

Agency: BROWARD COUNTY SO
Time of Inspection: 08:42

Date of Inspection: 04/18/2017

Serial Number: 80-001047
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.050	0.082	0.201	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.050	0.081	0.201	0.079

Standard Deviations	0.0004	0.0004	0.0007	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

04/18/2017
Date