

RECEIVED

Florida Department of Law Enforcement AUG 21 2017

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LARGO PD
Time of Inspection: 14:26

Date of Inspection: 08/21/2017

Serial Number: 80-001079
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check, Minimum Sample Volume Check, Alcohol Free Subject Test, Interferent Detect Test, and Diagnostic Check (Post-Inspection).

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for multiple samples.

Table with 5 columns: Standard Deviations, 0.0000, 0.0004, 0.0004, 0.0005.

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

Handwritten initials 'SP' in blue ink.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature in blue ink.

DANIELLE M BELL

Signature and Printed Name

08/21/2017
Date

Handwritten date '8/21/17' and initials in blue ink.

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

FEB 15 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LARGO PD
Time of Inspection: 13:50

Date of Inspection: 02/15/2017

Serial Number: 80-001079
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.076 | 0.195 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.080 |
| 0.000 | 0.048 | 0.078 | 0.197 | 0.079 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.197 | 0.079 |
| 0.000 | 0.049 | 0.078 | 0.197 | 0.079 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0007 | 0.0009 | 0.0004 |
|---------------------|--------|--------|--------|--------|


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DANIELLE M BELL

Signature and Printed Name

02/15/2017
Date