

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH SHORES
Time of Inspection: 14:16

Date of Inspection: 11/09/2017

Serial Number: 80-001242
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.049	0.079	0.201	0.077
0.000	0.049	0.080	0.202	0.077
0.000	0.049	0.080	0.202	0.078
0.000	0.050	0.080	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.049	0.079	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.050	0.080	0.201	0.077
0.000	0.050	0.080	0.201	0.077
Standard Deviations	0.0004	0.0004	0.0004	0.0003


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


PATRICK J MURPHY

 Signature and Printed Name

11/09/2017
 Date

11/9/17
20