

Florida Department of Law Enforcement Alcohol Testing Program

AUG 16 2017
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY S.O.
Time of Inspection: 13:36

Date of Inspection: 08/07/2017

Serial Number: 80-001254
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.078	0.195	0.081
0.000	0.049	0.079	0.196	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.199	0.080


Standard Deviations	0.0005	0.0006	0.0012	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

08/07/2017
Date

8/16/17
J/D

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 10 2017 *bc*

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BK

Agency: MONROE COUNTY S.O.
Time of Inspection: 12:27

Date of Inspection: 01/25/2017

Serial Number: 80-001254
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.050	0.078	0.197	0.080
0.000	0.051	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.051	0.079	0.199	0.081
0.000	0.051	0.079	0.197	0.081
0.000	0.051	0.080	0.198	0.080

Standard Deviations	0.0005	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

Dom

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

 Signature and Printed Name

01/25/2017
Date