

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 25 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.
Time of Inspection: 09:44

Date of Inspection: 10/25/2017

Serial Number: 80-001271
Software: 8100.27

BLC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.049	0.081	0.203	0.000 / 0.079
0.000	0.050	0.080	0.204	0.000 / 0.080
0.000	0.050	0.081	0.205	0.000 / 0.080
0.000	0.050	0.081	0.205	0.000 / 0.080
0.000	0.049	0.081	0.205	0.000 / 0.080
0.000	0.050	0.081	0.205	0.000 / 0.080
0.000	0.050	0.081	0.205	0.000 / 0.080
0.000	0.051	0.081	0.204	0.000 / 0.080
0.000	0.050	0.081	0.204	0.000 / 0.080
0.000	0.050	0.082	0.204	0.000 / 0.080

Standard Deviations	0.0005	0.0004	0.0006	0.0000 / 0.0003
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance DGS NOT CONNECTED.

pgm

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

10/25/2017
Date

*10/25/17
22*