

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NASSAU COUNTY SO  
Time of Inspection: 14:36

Date of Inspection: 08/30/2017

Serial Number: 80-001281  
Software: 8100.27

| Check or Test                                  | YES | NO | Check or Test                             | YES | NO |
|--|-----|----|---|-----|----|
| Diagnostic Check<br>(Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted                 |     | No |
| Minimum Sample Volume<br>Check: OK             | Yes |    | Barometric Pressure Sensor<br>Check: OK   | Yes |    |
| Alcohol Free Subject<br>Test: 0.000            | Yes |    | Mouth Alcohol Test:<br>Slope Not Met      | Yes |    |
| Interferent Detect Test:<br>Interferent Detect | Yes |    | Diagnostic Check<br>(Post-Inspection): OK | Yes |    |

| Alcohol Free Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#:201707D<br>Exp: 07/25/2019 | 0.08g/210L Test<br>(g/210L)<br>Lot#:201707E<br>Exp: 07/25/2019 | 0.20g/210L Test<br>(g/210L)<br>Lot#:201707C<br>Exp: 07/24/2019 | 0.08 g/210L<br>Dry Gas Std Test<br>(g/210L)<br>Lot#:AG619605<br>Exp: 07/14/2018 |
|-------------------------------|--|--|--|---|
| 0.000                         | 0.049  | 0.081  | 0.201  | 0.080   |
| 0.000                         | 0.049  | 0.081  | 0.201  | 0.079   |
| 0.000                         | 0.050  | 0.082  | 0.203  | 0.079   |
| 0.000                         | 0.051  | 0.082  | 0.202  | 0.079   |
| 0.000                         | 0.049  | 0.082  | 0.202  | 0.079   |
| 0.000                         | 0.050  | 0.083  | 0.202  | 0.079   |
| 0.000                         | 0.050  | 0.082  | 0.201  | 0.080   |
| 0.000                         | 0.049  | 0.082  | 0.201  | 0.080   |
| 0.000                         | 0.050  | 0.082  | 0.202  | 0.080   |
| 0.000                         | 0.049  | 0.082  | 0.201  | 0.080   |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0006 | 0.0005 | 0.0006 | 0.0005 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

PJM

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

08/30/2017  
Date

8/30/17  
JD