

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 27 2017

FDLE

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: OCALA POLICE DEPT
Time of Inspection: 15:46

Date of Inspection: 07/26/2017

Serial Number: 80-005454
Software: 8100.27

VBC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018
0.000	0.049	0.080	0.200	0.000 / 0.079
0.000	0.050	0.080	0.202	0.000 / 0.079
0.000	0.050	0.080	0.202	0.000 / 0.079
0.000	0.050	0.079	0.202	0.077 / 0.079
0.000	0.050	0.080	0.202	0.078 / 0.079
0.000	0.050	0.080	0.202	0.078 / 0.079
0.000	0.050	0.080	0.201	0.077 / 0.079
0.000	0.050	0.081	0.202	0.078 / 0.078
0.000	0.050	0.080	0.202	0.078 / 0.079
0.000	0.050	0.081	0.201	0.078 / 0.078

Standard Deviations	0.0003	0.0005	0.0006	0.0375 / 0.0004
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance DID NOT HAVE DGS CONNECTED

JLS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JD

JAKE L SHANAHAN
Signature and Printed Name

07/26/2017
Date