RECEIVED
AUG 02 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 12:46
Date of Inspection: 08/02/2017
Serial Number: 80-006027
Software: 8100.27

VBSC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.081	0.202	0.080
0.000	0.050	0.081	0.201	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
Standard Deviations	0.0000	0.0006	0.0016	0.0003

PSM

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt
Signature and Printed Name SHAYLA D PLATT

08/02/2017
Date

SP

FDLE/ATP Form 41 –Revised August 2005

RECEIVED
APR 04 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 12:05
Date of Inspection: 04/04/2017
Serial Number: 80-006027
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG612405 Exp: 05/03/2018
0.000	0.051	0.081	0.201	0.081
0.000	0.051	0.081	0.201	0.080
0.000	0.052	0.081	0.201	0.080
0.000	0.051	0.080	0.202	0.079
0.000	0.051	0.080	0.202	0.080
0.000	0.052	0.080	0.201	0.080
0.000	0.052	0.081	0.202	0.080
0.000	0.052	0.081	0.201	0.080
0.000	0.051	0.081	0.202	0.080
0.000	0.052	0.081	0.202	0.080
Standard Deviations	0.0005	0.0004	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt SHAYLA D PLATT
Signature and Printed Name

04/04/2017
Date

FDLE/ATP Form 41 –Revised August 2005

RECEIVED
FEB 10 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 11:47
Date of Inspection: 02/06/2017
Serial Number: 80-006027
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.050	0.080	0.202	0.080
0.000	0.051	0.080	0.202	0.080
0.000	0.051	0.080	0.203	0.081
0.000	0.052	0.081	0.203	0.081
0.000	0.051	0.080	0.203	0.080
0.000	0.051	0.080	0.203	0.081
0.000	0.052	0.081	0.204	0.081
0.000	0.052	0.081	0.204	0.080
0.000	0.052	0.081	0.204	0.080
0.000	0.052	0.081	0.204	0.080
Standard Deviations	0.0006	0.0005	0.0007	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt SHAYLA D PLATT
Signature and Printed Name

02/06/2017
Date

FDLE/ATP Form 41 –Revised August 2005