

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

MAR 13 2017

FDLE *WCC*  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 10:37

Date of Inspection: 03/09/2017

Serial Number: 80-006770  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: \_\_\_\_\_ Number of Simulators Used: \_\_\_\_\_

Remarks:  
ABORTED DI TO CHECK BREATH TUBE Non-compliance: .

*BK*

*Q*

**N/A COMPLIANCE  
NOT DETERMINED**

The above instrument complies ( ) does not comply (  ) with Chapter 11D-8, FAC.  
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*  
SHAYLA D PLATT  
Signature and Printed Name

03/09/2017  
Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAR 13 2017

FDLE *BCC*  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 15:44

Date of Inspection: 03/09/2017

Serial Number: 80-006770 *BK*  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.045	0.075	0.190	0.080
0.000	0.046	0.077	0.194	0.080
0.000	0.047	0.078	0.195	0.080
0.000	0.047	0.078	0.195	0.080
0.000	0.047	0.077	0.194	0.080
0.000	0.047	0.077	0.194	0.080
0.000	0.047	0.077	0.194	0.080
0.000	0.047	0.077	0.194	0.080
0.000	0.047	0.077	0.194	0.079
0.000	0.047	0.077	0.195	0.080
0.000	0.047	0.077	0.194	0.079

Standard Deviations	0.0006	0.0008	0.0014	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

*(Handwritten mark)*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

03/09/2017  
Date