



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency FDLE Evidential S/N 80-000206  
 Date In \_\_\_\_\_ Date Out 2/2/17  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>MS</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input checked="" type="checkbox"/> Other <u>Loaner</u> Visual Inspection: Case _____ Handle _____ Dry Gas Holder _____ Feet _____ Keyboard/Plug _____ Back/Plugs _____ Screws tight _____ Breath Hose _____ Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>MS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>226</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.156</u> (.139 - .169) 36mm <u>0.171</u> (.156 - .190) 53mm <u>0.238</u> (.228 - .278) 103mm <u>0.507</u> (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G11739</u></td> <td><u>201603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>G8149</u></td> <td><u>201601F</u> <u>11/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>G11621</u></td> <td><u>201604C</u> <u>4/15/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>A61619605</u> <u>7/14/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>G11739</u>	<u>201603D</u> <u>3/8/18</u>	0.08	<u>G8149</u>	<u>201601F</u> <u>11/26/18</u>	0.20	<u>G11621</u>	<u>201604C</u> <u>4/15/18</u>	0.08 DGS	N/A	<u>A61619605</u> <u>7/14/18</u>	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Program Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	<u>G11739</u>	<u>201603D</u> <u>3/8/18</u>															
0.08	<u>G8149</u>	<u>201601F</u> <u>11/26/18</u>															
0.20	<u>G11621</u>	<u>201604C</u> <u>4/15/18</u>															
0.08 DGS	N/A	<u>A61619605</u> <u>7/14/18</u>															

RECEIVED  
 FEB 03 2017  
 FDLE  
 Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			<b>Department Inspection</b> Performed By <u>MS</u> <input checked="" type="checkbox"/> Barometric Pressure _____ Gauge _____ ID# <u>26932</u> _____ Instrument _____ Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>G2879</u></td> </tr> <tr> <td>Interferent</td> <td><u>G8144</u></td> </tr> <tr> <td>0.05</td> <td><u>G11739</u></td> </tr> <tr> <td>0.08</td> <td><u>G8149</u></td> </tr> <tr> <td>0.20</td> <td><u>G11621</u></td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>Form 40</u>	Simulator	Serial Number	0.00	<u>G2879</u>	Interferent	<u>G8144</u>	0.05	<u>G11739</u>	0.08	<u>G8149</u>	0.20	<u>G11621</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	<u>G2879</u>																																																												
Interferent	<u>G8144</u>																																																												
0.05	<u>G11739</u>																																																												
0.08	<u>G8149</u>																																																												
0.20	<u>G11621</u>																																																												

Notes: QA/pc ok @pm 2/3/2017  
Shipping to Fort Walton Beach P.D. as  
evidential loaner GPM

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

B. Brett Kirkland

Quality Control Review

2/3/17  
Date

Stability Checks # 80-000206 FDLE 2/2/17 *QWS*

FDLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000206  
02/02/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:56
Control Test	0.047	11:57
Air Blank	0.000	11:58
Control Test	0.048	11:58
Air Blank	0.000	11:59
Control Test	0.047	11:59
Air Blank	0.000	12:00
Control Test Stats		
Average	0.0473	
Std Dev	0.0016	
Rel Std Dev(%)	1.2198	

*QWS*  
Operator's Signature

FDLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000206  
02/02/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:08
Control Test	0.076	12:08
Air Blank	0.000	12:09
Control Test	0.077	12:10
Air Blank	0.000	12:10
Control Test	0.078	12:11
Air Blank	0.000	12:11
Control Test Stats		
Average	0.0770	
Std Dev	0.0010	
Rel Std Dev(%)	1.2987	

*QWS*  
Operator's Signature

*Cleared  
condensation  
from tubing*  
*QWS*

FDLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000206  
02/02/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:45
Control Test	0.078	12:45
Air Blank	0.000	12:46
Control Test	0.078	12:47
Air Blank	0.000	12:47
Control Test	0.077	12:48
Air Blank	0.000	12:48
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

*QWS*  
Operator's Signature

FDLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000206  
02/02/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:13
Control Test	0.197	12:13
Air Blank	0.000	12:14
Control Test	0.201	12:15
Air Blank	0.000	12:15
Control Test	0.202	12:16
Air Blank	0.000	12:16
Control Test Stats		
Average	0.2000	
Std Dev	0.0026	
Rel Std Dev(%)	1.3229	

*QWS*  
Operator's Signature

*QWS*

FDLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000206  
02/02/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:31
Control Test	0.079	12:31
Air Blank	0.000	12:32
Control Test	0.079	12:32
Air Blank	0.000	12:32
Control Test	0.080	12:33
Air Blank	0.000	12:33
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

*QWS*  
*RSK*

*QWS*  
Operator's Signature