

INSTRUMENT PROCESSING SHEET

Agency Eschambia County S/N 80-000744

Date In 5/25/17 Date Out 5/30/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Lab Temp °C 20.6
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 235
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm .152 (.139 - .169)
 36mm .167 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .492 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>
0.20	<u>G11621</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 50mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By [Signature]

Simulator Temperatures °C
 External Digital Therm. ID#: 200505
 34°C +/- .2 Serial #: DR2035
 34°C +/- .2 Serial #: SD3964
 34°C +/- .2 Serial #: G11621

RECEIVED
JUN 02 2017
FDLE
Alcohol Testing Program

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1018 Gauge
 ID# 26932 1011 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2879</u>
Interferent	<u>G8144</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>G11621</u>

Attachments

Form 41 Calibration Adjustment
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes/Suggested Service:
QA/QC

[Signature] 6/2/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000744
 05/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.049	10:13
Air Blank	0.000	10:14
Control Test	0.049	10:14
Air Blank	0.000	10:15
Control Test	0.050	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	


 Operator's Signature

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000744
 05/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:07
Control Test	0.080	10:08
Air Blank	0.000	10:08
Control Test	0.081	10:09
Air Blank	0.000	10:09
Control Test	0.081	10:10
Air Blank	0.000	10:11
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	


 Operator's Signature

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000744
 05/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.198	09:59
Air Blank	0.000	10:00
Control Test	0.198	10:00
Air Blank	0.000	10:01
Control Test	0.200	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.1987	
Std Dev	0.0012	
Rel Std Dev(%)	0.5812	


 Operator's Signature

DGS

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000744
 05/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:27
Control Test	0.080	10:27
Air Blank	0.000	10:28
Control Test	0.079	10:28
Air Blank	0.000	10:29
Control Test	0.079	10:29
Air Blank	0.000	10:30
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS


 Operator's Signature

JE