



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jackson County SC S/N 80-000759
 Date In 10/10/17 Date Out 10/13/17 Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>PO9M</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.2</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>213</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.164</u> (.139 - .169) 36mm <u>.179</u> (.156 - .190) 53mm <u>.253</u> (.228 - .278) 103mm <u>.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G2835</u></td> <td><u>201702D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG708807</u> <u>3/29/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>G2835</u>	<u>201702D</u> <u>7/25/19</u>	0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.20	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.08 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>G2835</u>	<u>201702D</u> <u>7/25/19</u>															
0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>															
0.20	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>															
0.08 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By <u>PO9M</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD1025</u>															

RECEIVED
OCT 16 2017
Alcohol Testing Program

Calibration Adjustment Performed By _____																							
<input checked="" type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.300																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>PO9M</u>	
<input checked="" type="checkbox"/> Barometric Pressure <u>1013</u> Gauge ID# <u>28421</u> <u>1013</u> Instrument Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u>	
Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2840</u>
0.05	<u>G2835</u>
0.08	<u>DR1279</u>
0.20	<u>SD1025</u>
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	

Notes/Suggested Service:

QC 10/16/17
JJ Durham
 Quality Control Review

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

10/16/17
Date

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:26
Control Test	0.048	07:26
Air Blank	0.000	07:27
Control Test	0.047	07:28
Air Blank	0.000	07:28
Control Test	0.047	07:29
Air Blank	0.000	07:29
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

P Murphy
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:21
Control Test	0.077	07:21
Air Blank	0.000	07:22
Control Test	0.078	07:22
Air Blank	0.000	07:23
Control Test	0.078	07:24
Air Blank	0.000	07:24
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

P Murphy
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:16
Control Test	0.198	07:16
Air Blank	0.000	07:17
Control Test	0.198	07:17
Air Blank	0.000	07:18
Control Test	0.198	07:19
Air Blank	0.000	07:19
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000759
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:50
Control Test	0.078	07:50
Air Blank	0.000	07:51
Control Test	0.078	07:51
Air Blank	0.000	07:52
Control Test	0.078	07:52
Air Blank	0.000	07:53
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P Murphy
Operator's Signature

10/16/17
JD