

INSTRUMENT PROCESSING SHEET

Agency Alachua County

S/N 80-000760

Date In 2/15/17

Date Out 3/2/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 175
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 155 (.139 - .169)
 36mm 170 (.156 - .190)
 53mm 241 (.228 - .278)
 103mm 502 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	201603D 3/8/18
0.08	SD1011	201601F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG626604 9/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 417mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAR 02 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure ID# 26932 Gauge 1016 Instrument 1014

Mouth Alcohol Solution Lot # 2017-A 2016A
 Acetone Stock Solution Lot # 2016-A 2017A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	DR2035
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: paper feed button will not advance paper. Still prints when prompted to though. AZ can determine if they want it sent to repair.

DA/PC OK BDM 3/2/2017
Travis Kankler

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date 3/2/17

80-000760
 Stability checks
 3/2/17

INTOXILYZER 8000
 Instrument Initialization
 06:08 03/02/2017

ALACHUA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000760
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:56
Control Test	0.079	06:56
Air Blank	0.000	06:57
Control Test	0.079	06:57
Air Blank	0.000	06:58
Control Test	0.078	06:58
Air Blank	0.000	06:59
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DS

Operator's Signature

ALACHUA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000760
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:00
Control Test	0.049	07:01
Air Blank	0.000	07:01
Control Test	0.049	07:02
Air Blank	0.000	07:02
Control Test	0.049	07:03
Air Blank	0.000	07:04
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

ALACHUA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000760
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:06
Control Test	0.197	07:06
Air Blank	0.000	07:07
Control Test	0.197	07:07
Air Blank	0.000	07:08
Control Test	0.198	07:09
Air Blank	0.000	07:09
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

Operator's Signature

ALACHUA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000760
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:12
Control Test	0.079	07:13
Air Blank	0.000	07:13
Control Test	0.080	07:14
Air Blank	0.000	07:14
Control Test	0.080	07:15
Air Blank	0.000	07:16
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DS
 YSK

Operator's Signature