



INSTRUMENT PROCESSING SHEET

Agency Okaloosa County SOS/N 80-000763

Florida Department of Law Enforcement

Date In 10/30/2017 DI Completion Date 11/3/17 Ship P/U H/D CMI EE

Intake Performed By <u>ATP</u>	Quality Checks Performed By <u>BSM</u>	Flow Calibration Performed By _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>215</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)

Final Release Date

NOV 06 2017

Simulator	Serial #	Lot #/Exp
0.050	<u>G2835</u>	201707D 07/25/2019
0.080	<u>DR1279</u>	201707E 07/25/2019
0.200	<u>SD1025</u>	201707C 07/24/2019
0.080 DGS	N/A	<u>AG626605</u> <u>9/22/18</u>

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By BSM

Lab Temp °C 22.3
 External Digital Therm. ID#: 800504
 34°C +/- .2 Serial #: G2835
 34°C +/- .2 Serial #: DR1279
 34°C +/- .2 Serial #: SD1025

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By BSM

Barometric Pressure ID# 28421
 Gauge 1019 Instrument 1019
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	<u>G2880</u>
Interferent	<u>G2840</u>
0.050	<u>G2835</u>
0.080	<u>DR1279</u>
0.200	<u>SD1025</u>

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other _____
 Post-Stability Checks

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

BSM 11/6/17 J. Bohan 11/6/17
 Tech Review / Date Admin Review / Date

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000763
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:20
Control Test	0.050	12:20
Air Blank	0.000	12:21
Control Test	0.050	12:21
Air Blank	0.000	12:22
Control Test	0.050	12:23
Air Blank	0.000	12:23
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000763
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:14
Control Test	0.081	12:15
Air Blank	0.000	12:15
Control Test	0.082	12:16
Air Blank	0.000	12:17
Control Test	0.081	12:17
Air Blank	0.000	12:18
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

P Murphy
Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000763
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:29
Control Test	0.204	12:30
Air Blank	0.000	12:30
Control Test	0.201	12:31
Air Blank	0.000	12:32
Control Test	0.201	12:32
Air Blank	0.000	12:33
Control Test Stats		
Average	0.2020	
Std Dev	0.0017	
Rel Std Dev(%)	0.8574	

P Murphy
Operator's Signature

DB

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000763
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:52
Control Test	0.082	12:53
Air Blank	0.000	12:53
Control Test	0.082	12:54
Air Blank	0.000	12:54
Control Test	0.081	12:54
Air Blank	0.000	12:55
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

DGS

P Murphy
Operator's Signature

11/6/17
JD