



INSTRUMENT PROCESSING SHEET

Agency Hamilton CountyS/N 80-000797

Florida Department of Law Enforcement

Date In 12/5/17DI Completion Date 12/6/17 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>200</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>100</u> (.139 - .169) 36 mm <u>181</u> (.156 - .190) 53 mm <u>250</u> (.228 - .278) 103 mm <u>523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	--	--

Final Release Date FDLE DEC 08 2017 Alcohol Testing Program	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>62835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG715202</u> <u>6/1/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.3</u> External Digital Therm. ID#: <u>30504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>
Simulator	Serial #	Lot #/Exp															
0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>															
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>															
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>															
0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>															

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>[Signature]</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1016</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2840</u></td> </tr> <tr> <td>0.050</td> <td><u>G2835</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2880</u>	Interferent	<u>G2840</u>	0.050	<u>G2835</u>	0.080	<u>DR1279</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	<u>G2880</u>																																																												
Interferent	<u>G2840</u>																																																												
0.050	<u>G2835</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>SD1025</u>																																																												

Notes/Suggested Service: <u>AZ needs to change level 2 Pw to smethy unipe</u> <u>[Signature]</u> _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>[Signature]</u> <u>12/6/17</u> <u>[Signature]</u> <u>12/8/17</u> Tech Review / Date Admin Review / Date
--	--

80-000797

Stability Access
12/6/17 @

INTOXILYZER 8000
Instrument Initialization
10:28 12/06/2017

HAMILTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000797
12/06/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:14
Control Test	0.079	11:15
Air Blank	0.000	11:15
Control Test	0.078	11:16
Air Blank	0.000	11:16
Control Test	0.079	11:17
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

HAMILTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000797
12/06/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:09
Control Test	0.200	11:10
Air Blank	0.000	11:10
Control Test	0.200	11:11
Air Blank	0.000	11:11
Control Test	0.201	11:12
Air Blank	0.000	11:13
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

HAMILTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000797
12/06/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:05
Control Test	0.081	11:05
Air Blank	0.000	11:06
Control Test	0.081	11:06
Air Blank	0.000	11:06
Control Test	0.081	11:07
Air Blank	0.000	11:07
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

HAMILTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000797
12/06/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:19
Control Test	0.049	11:20
Air Blank	0.000	11:21
Control Test	0.049	11:21
Air Blank	0.000	11:22
Control Test	0.049	11:23
Air Blank	0.000	11:23
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

12/8/17
JA