FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Dixie County SO S/N 80-000809

Date In 2/22/17 Date Out 3/2/17 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED
MAR 02 2017
Alcohol Testing Program
FDLE

Intake Performed By TP

- ☐ Registration
- ☑ Annual
- ☑ Return from CMI
- ☐ Return from Enforcement
- ☐ Electronics
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:
- ☑ Power cord
- ☐ Printer Cable
- ☐ Other Static Bag

Notes:

Quality Checks Performed By

- ☑ Breath Tube Screen
- ☐ Replace O-Rings
- ☐ Instrument Set Up Verified
- ☑ R-Value 202
- ☑ Flow Verification (L/s)

Flow Column # ATP105

32mm	157	(.139 - .169)
36mm	176	(.156 - .190)
53mm	243	(.228 - .278)
103mm	504	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 26932

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	201603D
		3/8/18
0.08	SD1011	201601P
		1/26/18
0.20	SD1025	201604C
		4/5/18
0.08 DGS	N/A	AG6266604
		9/22/18

Flow Calibration Performed By

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column #
- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32mm	(.139 - .169)
36mm	(.156 - .190)
53mm	(.228 - .278)
103mm	(.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

- ☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By

- ☑ Barometric Pressure ID# 26932 — 1016 Gauge / 1016 Instrument

Mouth Alcohol Solution Lot # 2016-A

Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	DR2035
0.08	SD1011
0.20	SD1025

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: DGS Not attached during first run of 08 gas on Inspection. Attached and retested. In Compliance.

DA/QC OK PSM 3/2/2017

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

Brett Strickland
Quality Control Review

3/2/17
Date

Revised November 2012

80-000809
Stability Checks
3/2/17

INTOXILYZER 8000
Instrument Initialization
06:11 03/02/2017

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:56
Control Test	0.079	06:56
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:58
Control Test	0.080	06:58
Air Blank	0.000	06:59
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DRS

Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:00
Control Test	0.198	07:01
Air Blank	0.000	07:01
Control Test	0.199	07:02
Air Blank	0.000	07:03
Control Test	0.200	07:05
Air Blank	0.000	07:04
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:05
Control Test	0.079	07:06
Air Blank	0.000	07:06
Control Test	0.080	07:07
Air Blank	0.000	07:08
Control Test	0.080	07:08
Air Blank	0.000	07:09
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:12
Control Test	0.049	07:12
Air Blank	0.000	07:13
Control Test	0.048	07:14
Air Blank	0.000	07:14
Control Test	0.049	07:15
Air Blank	0.000	07:15
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

BSM

BK

Operator's Signature

FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Dixie County S/N 80-000809

Date In 1/25/17 Date Out 2/8/17 ☑Ship ☐P/U ☐H/D ☑CMI ☐EE

RECEIVED
FEB 10 2017
FDLE
Alcohol Testing Program

Intake Performed By [illegible]

- ☐ Registration
- ☑ Annual
- ☑ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other

Visual Inspection:

ok Case	ok Handle
ok Dry Gas Holder	ok Feet
ok Keyboard/Plug	ok Back/Plugs
ok Screws tight	ok Breath Hose

Other Equipment:

- ☑ Power cord
- ☐ Printer Cable
- ☑ Other Static Bag

Notes:

Quality Checks Performed By [illegible]

- ☑ Breath Tube Screen
- ☑ Replace O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 200
- ☑ Flow Verification (L/s)

Flow Column # ATP102

32mm 151 (.139 - .169)
36mm 177 (.156 - .190)
53mm 264 (.228 - .278)
103mm 507 (.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28421

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	20160 3D 3/8/18
0.08	G8149	20160 1F 1/26/18
0.20	G11621	20160 4C 4/5/18
0.08 DGS	N/A	AGG 266 05 9/22/18

Flow Calibration Performed By

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column #

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☑ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

- ☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [illegible]

- ☑ Barometric Pressure 1009 Gauge
  ID# 28421 1010 Instrument

Mouth Alcohol Solution Lot # 2016-C
Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8149
0.20	G11621

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: Printer communication issue. Will not print to multiple external printers. Will only attempt to print to internal and error out. Recommending it being sent back to CMI since it just came back from there.

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☑ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

QC KDVS
Brett Burkland
Quality Control Review

2/10/17
Date

Revised November 2012

80-000809
Stab. 11 my checks
2/8/17
RK

INTOXILYZER 8000
Instrument Initialization
07:18 02/08/2017

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
02/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:47
Control Test	0.080	08:48
Air Blank	0.000	08:48
Control Test	0.080	08:48
Air Blank	0.000	08:49
Control Test	0.080	08:49
Air Blank	0.000	08:50
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
02/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.048	08:57
Air Blank	0.000	08:57
Control Test	0.049	08:58
Air Blank	0.000	08:59
Control Test	0.049	08:59
Air Blank	0.000	09:00
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
02/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.079	09:02
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:05
Air Blank	0.000	09:05
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000809
02/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.199	09:07
Air Blank	0.000	09:08
Control Test	0.200	09:08
Air Blank	0.000	09:09
Control Test	0.200	09:09
Air Blank	0.000	09:10
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

Operator's Signature