

Stability Checks # 80-000812 Hernando County S.O. 10/18/17 ~~DES~~

DES

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
10/18/2017
Software: 8100.27

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
10/18/2017
Software: 8100.27

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
10/18/2017
Software: 8100.27

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
10/18/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:41
Control Test	0.081	10:41
Air Blank	0.000	10:42
Control Test	0.081	10:42
Air Blank	0.000	10:42
Control Test	0.081	10:42
Air Blank	0.000	10:43
Control Test	0.000	10:43
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	10:36
Control Test	0.199	10:37
Air Blank	0.000	10:37
Control Test	0.201	10:38
Air Blank	0.000	10:39
Control Test	0.201	10:39
Air Blank	0.000	10:40
Control Test	0.000	10:40
Average	0.2003	
Std Dev	0.0012	
Rel Std Dev(%)	0.5764	

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.081	10:32
Air Blank	0.000	10:32
Control Test	0.083	10:33
Air Blank	0.000	10:34
Control Test	0.082	10:34
Air Blank	0.000	10:35
Control Test	0.0820	10:35
Average	0.0820	
Std Dev	0.0010	
Rel Std Dev(%)	1.2195	

Test	g/210L	Time
Air Blank	0.000	10:26
Control Test	0.051	10:27
Air Blank	0.000	10:28
Control Test	0.051	10:28
Air Blank	0.000	10:29
Control Test	0.050	10:30
Air Blank	0.000	10:30
Control Test	0.0507	10:30
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

DES
Operator's Signature

DES
Operator's Signature

DES
Operator's Signature

DES
Operator's Signature

10/18/17
DES