



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

RECEIVED

Agency Marion County S/N 80-000827 MAR 28 2017

Date In 3/23/17 Date Out 3/28/17 Ship P/U H/D CMI FDEE

Alcohol Testing Program

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ck Case ck Handle
ck Dry Gas Holder ck Feet
ck Keyboard/Plug ck Back/Plugs
ck Screws tight ck Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 157
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 143 (.139 - .169)
 36mm 161 (.156 - .190)
 53mm 240 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD3964</u>	<u>201601P</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure ID# 28421 Gauge 1014 Instrument 1014

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>SD1025</u>

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: Change level 2 PW to unique PW.
QA/AC OK QDM 3/28/2017 [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]

Quality Control Review

3/28/17
Date

80-000887
 Stability Checks
 3/28/17

INTOXILYZER 8000
 Instrument Initialization
 07:29 03/28/2017

MARION COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000827
 03/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:06
Control Test	0.049	08:07
Air Blank	0.000	08:08
Control Test	0.050	08:08
Air Blank	0.000	08:09
Control Test	0.050	08:10
Air Blank	0.000	08:10
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1625	

Operator's Signature

DS

MARION COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000827
 03/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:56
Control Test	0.078	07:57
Air Blank	0.000	07:57
Control Test	0.078	07:57
Air Blank	0.000	07:58
Control Test	0.078	07:58
Air Blank	0.000	07:59
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

Operator's Signature

MARION COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000827
 03/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:02
Control Test	0.080	08:02
Air Blank	0.000	08:03
Control Test	0.081	08:03
Air Blank	0.000	08:04
Control Test	0.079	08:05
Air Blank	0.000	08:05
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel. Std Dev(%)	1.2500	

Operator's Signature

MARION COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000827
 03/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:16
Control Test	0.199	08:17
Air Blank	0.000	08:17
Control Test	0.200	08:18
Air Blank	0.000	08:18
Control Test	0.201	08:19
Air Blank	0.000	08:20
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel. Std Dev(%)	0.5000	

Operator's Signature

PDM
BK