

INSTRUMENT PROCESSING SHEET

Agency Hillsborough County S/N 80-000833
 Date In 2/13/17 Date Out 3/31/17 Ship P/U H/D CMI EE

Intake Performed By DB

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: Spare Parts included

Quality Checks Performed By DBS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 195
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 0.156 (.139 - .169)
 36mm 0.171 (.156 - .190)
 53mm 0.242 (.228 - .278)
 103mm 0.519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	<u>N/A</u>	<u>AG626605</u> <u>9/22/18</u>

Flow Calibration Performed By DBS

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAR 06 2017
 FDLE
 Alcohol Testing Program

ID# 28421
 #2
 1028
 1031

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DBS

Barometric Pressure 1029 Gauge
 ID# 28421 1032 Instrument

Mouth Alcohol Solution Lot # 2016-C / 2016-C
 Acetone Stock Solution Lot # 2017-A / 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD1011</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other Attach Form 41

Notes: Initial Dept Insp. lost due to power failure DBS
DB @ 3/6/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

3/6/17
 Date

Stability Checks # 80-000833 Hillsborough County SO. 3/3/17 ~~008~~

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000833
03/03/2017
Software: 8100.27

Test	9/21:0L	Time
Air Blank	0.000	12:36
Control Test	0.050	12:37
Air Blank	0.000	12:37
Control Test	0.050	12:38
Air Blank	0.000	12:39
Control Test	0.050	12:39
Air Blank	0.000	12:40
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000833
03/03/2017
Software: 8100.27

Test	9/21:0L	Time
Air Blank	0.000	12:41
Control Test	0.081	12:42
Air Blank	0.000	12:42
Control Test	0.080	12:43
Air Blank	0.000	12:43
Control Test	0.081	12:44
Air Blank	0.000	12:45
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	


Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000833
03/03/2017
Software: 8100.27

Test	9/21:0L	Time
Air Blank	0.000	12:47
Control Test	0.196	12:47
Air Blank	0.000	12:48
Control Test	0.197	12:48
Air Blank	0.000	12:49
Control Test	0.197	12:50
Air Blank	0.000	12:50
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	


Operator's Signature

DGS

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000833
03/03/2017
Software: 8100.27

Test	9/21:0L	Time
Air Blank	0.000	12:52
Control Test	0.081	12:52
Air Blank	0.000	12:53
Control Test	0.080	12:53
Air Blank	0.000	12:54
Control Test	0.081	12:54
Air Blank	0.000	12:54
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	


Operator's Signature