FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Monroe County Sheriff's Office S/N 80-000869

Date In 5/3/2017 Date Out 5/3/2017 ☑ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By DERR
- ☐ Registration
- ☑ Annual
- ☐ Return from CMI
- ☑ Return from Enforcement Electronics
- ☐ Other ____

Visual Inspection:
- OK Case — OK Handle
- OK Dry Gas Holder — OK Feet
- OK Keyboard/Plug — OK Back/Plugs
- OK Screws tight — OK Breath Hose

Other Equipment:
- ☐ Power cord
- ☐ Printer Cable
- ☑ Other ANTI STATIC BAG

Notes: ____

Quality Checks Performed By DERR
- ☑ Breath Tube Screen
- ☑ Replace O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 220
- ☑ Flow Verification (L/s)

Flow Column # ATP 101
- 32mm .156 (.139 - .169)
- 36mm .175 (.156 - .190)
- 53mm .246 (.228 - .278)
- 103mm .503 (.447 - .547)

☑ Barometric Pressure Check
Gauge ID # 281663
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1014	201603D 03/08/2018
0.08	SD1015	201601F 01/26/2018
0.20	SD1017	201604C 04/05/2018
0.08 DGS	N/A	A6626605 9/22/2018

Flow Calibration Performed By ____
- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column # ____
- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value ____
- ☐ Post Calibration Verification (L/s)

Flow Column # ____
- 32mm ____ (.139 - .169)
- 36mm ____ (.156 - .190)
- 53mm ____ (.228 - .278)
- 103mm ____ (.447 - .547)

RECEIVED
MAY 09 2017
FDLE
Alcohol Testing Program

Maintenance Performed By ____
- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other ____

Suggested Service
____

Optical Bench Calibration Performed By ____
- ☑ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ____ ID # ____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DERR

☑ Barometric Pressure ID# 28199 — 1015 Gauge — 1015 Instrument

Mouth Alcohol Solution Lot # 2016A
Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	G8147
Interferent	G12100
0.05	SD1014
0.08	SD1015
0.20	SD1017

Attachments
- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other ____

Notes: E-MAILED 5/3/2017 APPROVED
QA/QC OK [initials] 5/9/2017

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place Into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

[signature] 5/9/17
Quality Control Review — Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000869	Monroe County Sheriff's Office	5/3/2017	

0.05g/210L
0.047 to 0.053 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
05/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:36
Control Test	0.051	10:37
Air Blank	0.000	10:37
Control Test	0.051	10:38
Air Blank	0.000	10:39
Control Test	0.051	10:39
Air Blank	0.000	10:40
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

0.08g/210L
0.077 to 0.083 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
05/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:41
Control Test	0.080	10:42
Air Blank	0.000	10:42
Control Test	0.081	10:43
Air Blank	0.000	10:44
Control Test	0.081	10:44
Air Blank	0.000	10:45
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Operator's Signature

0.20g/210L
0.194 to 0.206 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
05/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:48
Control Test	0.200	10:48
Air Blank	0.000	10:49
Control Test	0.201	10:50
Air Blank	0.000	10:50
Control Test	0.200	10:51
Air Blank	0.000	10:51
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

DGS 0.08g/210L
0.077 to 0.083 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
05/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:53
Control Test	0.080	10:53
Air Blank	0.000	10:54
Control Test	0.080	10:54
Air Blank	0.000	10:55
Control Test	0.081	10:55
Air Blank	0.000	10:56
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Monroe County Sheriffs Office S/N 80-000869

Date In 01/25/2017 Date Out 01/25/2017 ☐Ship ☑P/U ☐H/D ☐CMI ☐EE

RECEIVED
FEB 10 2017
FDLE
Alcohol Testing Program

Intake
Performed By DERR

- ☐ Registration
- ☑ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other ____

Visual Inspection:

OK Case
OK Handle
OK Dry Gas Holder
OK Feet
OK Keyboard/Plug
OK Back/Plugs
OK Screws tight
OK Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☐ Other ____

Notes: ____

Quality Checks
Performed By DERR

- ☑ Breath Tube Screen
- ☑ Replace O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 221
- ☑ Flow Verification (L/s)

Flow Column # ATP 101

32mm 156 (.139 - .169)
36mm 175 (.156 - .190)
53mm 246 (.228 - .278)
103mm 507 (.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28663

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD 3967	201603D 03/08/2018
0.08	SD 3968	201601F 01/26/2018
0.20	SD 3969	201604C 04/05/2018
0.08 DGS	N/A	AG605301 02/22/2018

Flow Calibration
Performed By ____

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column # ____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value ____
- ☐ Post Calibration Verification (L/s)

Flow Column # ____

32mm ____ (.139 - .169)
36mm ____ (.156 - .190)
53mm ____ (.228 - .278)
103mm ____ (.447 - .547)

Maintenance
Performed By ____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other ____

Suggested Service ____

Optical Bench Calibration
Performed By DERR

- ☑ Optical Bench Calibration N/A
- ☑ Optical Bench Calibration Complete

Barometric Pressure Gauge 1014 ID # 68639

Simulator	Serial Number	Lot Number	Expiration
0.000	2235	N/A	N/A
0.040	2236	16101	02-02-2018
0.100	2237	15001	05-20-2017
0.200	2238	15104	05-27-2017
0.400	2239	16102	03-22-2018
0.080 DGS	N/A	034150801A1	03-05-2017

- ☑ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	SD3967	201603D	03/08/2018
0.08	SD3968	201601F	01/26/2018
0.20	SD3969	201604C	04/05/2018
0.08 DGS	N/A	AG605301	02/22/2018

Department Inspection
Performed By DERR

- ☑ Barometric Pressure ID# 68639/28199 1016/1014 Gauge 1015/1012 Instrument

Mouth Alcohol Solution Lot # 2016A / 2016A
Acetone Stock Solution Lot # 2016B / 2016B

Simulator	Serial Number
0.00	SD 3965/ SD 3965
Interferent	SD 3966/ SD 3966
0.05	SD 3967/ SD 3967
0.08	SD 3968/ SD 3968
0.20	SD 3969/ SD 3969

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☑ Optical Bench Cal
- ☑ Post-Stability Tests
- ☑ Other Form 40

Notes: E-MAILED 01/25/2017
After non compliance DI I recalibrated instrument to bring values closer to nominal. Inst Failed DI AFTER calibration. Instrument will be sent to EE for repair.

- ☐ Instrument Complies with Chapter 11D-8, FAC
- ☑ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☐ Return to/Place into Evidentiary Use
- ☑ Remain Out of Evidentiary Use
- ☐ Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

2/10/17
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-000869	Monroe County Sheriff's Office	1/25/2017	

0.05g/210L

0.047 to 0.053 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:23
Control Test	0.049	13:24
Air Blank	0.000	13:24
Control Test	0.050	13:25
Air Blank	0.000	13:25
Control Test	0.048	13:26
Air Blank	0.000	13:27

Control Test Stats

Average	0.0490
Std Dev	0.0010
Rel Std Dev(%)	2.0408

Operator's Signature

0.08g/210L

0.077 to 0.083 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:28
Control Test	0.079	13:29
Air Blank	0.000	13:29
Control Test	0.079	13:30
Air Blank	0.000	13:30
Control Test	0.080	13:31
Air Blank	0.000	13:32

Control Test Stats

Average	0.0793
Std Dev	0.0006
Rel Std Dev(%)	0.7277

Operator's Signature

0.20g/210L

0.194 to 0.206 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:33
Control Test	0.196	13:33
Air Blank	0.000	13:34
Control Test	0.198	13:35
Air Blank	0.000	13:35
Control Test	0.198	13:36
Air Blank	0.000	13:37

Control Test Stats

Average	0.1973
Std Dev	0.0012
Rel Std Dev(%)	0.5852

Operator's Signature

DGS 0.08g/210L

0.077 to 0.083 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:38
Control Test	0.080	13:38
Air Blank	0.000	13:39
Control Test	0.080	13:39
Air Blank	0.000	13:40
Control Test	0.080	13:40
Air Blank	0.000	13:41

Control Test Stats

Average	0.0800
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Operator's Signature

```
MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000                SN 80-000869
01/25/2017                    12:39:35

Auto Calibration
Max Power Res Value = 35
Auto Range Res Value = 28

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12920, 9um Io = 14090
     <<<<<  CHANNEL 1  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  0.0840     (-0.0070)
Sample #2 =  0.0950     (0.0210)
Sample #3 =  0.0940     (0.1400)
Sample #4 =  0.1040     (0.1350)
Avg % Abs = 0.0977  (0.0987)
STD DEV =  0.0055  (0.0673)
REL STD DEV = 5.639  (68.217)
        -------------------

     <<<<<  CHANNEL 2  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  0.1000     (-0.0220)
Sample #2 =  0.0950     (-0.0080)
Sample #3 =  0.1030     (0.0480)
Sample #4 =  0.0900     (0.0590)
Avg % Abs = 0.0960  (0.0330)
STD DEV =  0.0066  (0.0359)
REL STD DEV = 6.831  (108.880)
        -------------------

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12890, 9um Io = 14077
     <<<<<  CHANNEL 1  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  0.9020     (-0.0180)
Sample #2 =  0.8150     (0.0770)
Sample #3 =  0.8400     (0.1210)
Sample #4 =  0.8220     (0.1550)
Avg % Abs = 0.8257  (0.1177)
STD DEV =  0.0129  (0.0391)
REL STD DEV = 1.562  (33.235)
        -------------------

     <<<<<  CHANNEL 2  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  1.4700     (0.0000)
Sample #2 =  1.4250     (0.0520)
Sample #3 =  1.4420     (0.0710)
Sample #4 =  1.4340     (0.0880)
Avg % Abs = 1.4337  (0.0703)
STD DEV =  0.0085  (0.0180)
REL STD DEV = 0.593  (25.606)
        -------------------

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12838, 9um Io = 14044
     <<<<<  CHANNEL 1  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  1.9630     (-0.0060)
Sample #2 =  1.9410     (0.0680)
Sample #3 =  1.9390     (0.1170)
Sample #4 =  1.9330     (0.2160)
Avg % Abs = 1.9377  (0.1337)
STD DEV =  0.0042  (0.0754)
REL STD DEV = 0.215  (56.405)
        -------------------

     <<<<<  CHANNEL 2  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  3.4980     (-0.0270)
Sample #2 =  3.4700     (0.0400)
Sample #3 =  3.4710     (0.0740)
Sample #4 =  3.4860     (0.1090)
Avg % Abs = 3.4757  (0.0743)
STD DEV =  0.0090  (0.0345)
REL STD DEV = 0.258  (46.414)
        -------------------

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12793, 9um Io = 14021
     <<<<<  CHANNEL 1  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  3.7520     (0.0150)
Sample #2 =  3.6920     (0.1660)
Sample #3 =  3.6620     (0.2220)
Sample #4 =  3.6640     (0.2570)
Avg % Abs = 3.6727  (0.218[illegible])
STD DEV =  0.0168  (0.0506)
REL STD DEV = 0.457  (23.175)
        -------------------

     <<<<<  CHANNEL 2  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  6.7150     (0.0000)
Sample #2 =  6.6540     (0.1140)
Sample #3 =  6.6570     (0.1390)
Sample #4 =  6.6570     (0.1680)
Avg % Abs = 6.6560  (0.1403)
STD DEV =  0.0017  (0.0270)
REL STD DEV = 0.026  (19.257)
        -------------------

Sol Value = 0.400 g/210L ***
Fit value = 1.9048 mg/l  %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12746, 9um Io = 13994
     <<<<<  CHANNEL 1  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 =  7.0550     (-0.0050)
Sample #2 =  7.0420     (0.1600)
Sample #3 =  6.9520     (0.2600)
Sample #4 =  6.9650     (0.3200)
Avg % Abs = 6.9863  (0.2467)
STD DEV =  0.0486  (0.0808)
REL STD DEV = 0.696  (32.769)
        -------------------

     <<<<<  CHANNEL 2  >>>>>
 Sample        % Abs     (% Abs Ref)
Sample #1 = 12.5780     (-0.0280)
Sample #2 = 12.5120     (0.1450)
Sample #3 = 12.4560     (0.1990)
Sample #4 = 12.4270     (0.2490)
Avg % Abs = 12.4650  (0.1977)
STD DEV =  0.0432  (0.0520)
REL STD DEV = 0.347  (26.313)
        -------------------
```

Optical Calibration
SN: 80-000869
Agency: Monroe County SO.
Date: 01/25/2017
Quadratic Fit: +/-0.002g/210L
By: [illegible]

```
   ***** AUTO CAL DATA *****
     <<<<<  CHANNEL 1  >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs =    0.098
  Std Dev =   0.01 Rel Std Dev =   5.64
Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs =    0.826
  Std Dev =   0.01 Rel Std Dev =   1.56
Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs =    1.938
  Std Dev =   0.00 Rel Std Dev =   0.21
Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs =    3.673
  Std Dev =   0.02 Rel Std Dev =   0.46
Sol Val = 1.9048 mg/l or 0.400 g/210L
 % Abs =    6.986
  Std Dev =   0.05 Rel Std Dev =   0.70
Zero Order Coef = -239.49
First Order Coef = 2535.86
Second Order Coef =  32.22
Standard Deviation =  23.598946
        -------------------

     <<<<<  CHANNEL 2  >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs =    0.096
  Std Dev =   0.01 Rel Std Dev =   6.83
Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs =    1.434
  Std Dev =   0.01 Rel Std Dev =   0.59
Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs =    3.476
  Std Dev =   0.01 Rel Std Dev =   0.26
Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs =    6.656
  Std Dev =   0.00 Rel Std Dev =   0.03
Sol Val = 1.9048 mg/l or 0.400 g/210L
 % Abs =   12.465
  Std Dev =   0.04 Rel Std Dev =   0.35
Zero Order Coef = [illegible]
First Order Coef = 1347.35
Second Order Coef =  15.11
Standard Deviation =  26.731756
        -------------------
```

Solution Stats Quadratic Fit Chan 1

Act g/210L	Fit g/210L	Residual g/210L
0.000	0.000	-0.0002
0.040	0.039	0.0006
0.100	0.101	-0.0007
0.200	0.200	0.0003
0.400	[illegible]	-0.0000

Solution Stats Quadratic Fit Chan 2

Act g/210L	Fit g/210L	Residual g/210L
0.000	0.001	-0.0006
0.040	0.039	0.0009
0.100	0.100	-0.0001
0.200	0.200	-0.0003
0.400	0.400	0.0001

```
Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l  %%%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 =  2955.00
Sample #2 =  3032.00
Sample #3 =  2958.00
Sample #4 =  3181.00
Average Result = 3057.0000
STD DEV = 113.5826
REL STD DEV = 3.715
**********
***** CHANNEL 2
Sample #1 =  3271.00
Sample #2 =  3294.00
Sample #3 =  3266.00
Sample #4 =  3378.00
Average Result = 3312.6667
STD DEV = 58.2866
REL STD DEV = 1.760
**********
Dry Gas H20 Adjust Results **********
  Barometric Pressure = 1013
  3 um H20 Adjust (mg/l*10,000) = 752
  9 um H20 Adjust (mg/l*10,000) = 497
**** AUTO CAL PASS
```

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000869	Monroe County Sheriff's Office	01/25/2017	

0.05g/210L
0.047 to 0.053 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	0.051	09:47
Air Blank	0.000	09:48
Control Test	0.050	09:49
Air Blank	0.000	09:49
Control Test	0.050	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Operator's Signature

0.08g/210L
0.077 to 0.083 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.081	09:52
Air Blank	0.000	09:53
Control Test	0.081	09:54
Air Blank	0.000	09:54
Control Test	0.081	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

0.20g/210L
0.194 to 0.206 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.201	09:58
Air Blank	0.000	09:58
Control Test	0.200	09:59
Air Blank	0.000	09:59
Control Test	0.199	10:00
Air Blank	0.000	10:01
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

Operator's Signature

DGS 0.08g/210L
0.077 to 0.083 ☑

MONROE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000869
01/25/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.078	10:02
Air Blank	0.000	10:03
Control Test	0.080	10:03
Air Blank	0.000	10:04
Control Test	0.080	10:04
Air Blank	0.000	10:05
Control Test Stats		
Average	0.0793	
Std Dev	0.0012	
Rel Std Dev(%)	1.4555	

Operator's Signature