FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Miami Police Department S/N 80-000874

Date In 6-7-2017 Date Out 6-12-2017 ☑ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED JUN 26 2017 FDLE Alcohol Testing Program

Intake Performed By DERR

- ☐ Registration
- ☑ Annual
- ☐ Return from CMI
- ☑ Return from Enforcement
- Electronics
- ☐ Other

Visual Inspection:

OK	Case	OK	Handle
OK	Dry Gas Holder	OK	Feet
OK	Keyboard/Plug	OK	Back/Plugs
OK	Screws tight	OK	Breath Hose

Other Equipment:
- ☐ Power cord
- ☐ Printer Cable
- ☐ Other

Notes:

Quality Checks Performed By DERR

- ☑ Lab Temp °C 22.34
- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 197
- ☑ Flow Verification (L/s)

Flow Column # ATP101
32mm .156 (.139 - .169)
36mm .175 (.156 - .190)
53mm .250 (.228 - .278)
103mm .511 (.447 - .547)

- ☑ Barometric Pressure Check
- Gauge ID # 28663
- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3967	201603D 03/08/2018
0.08	SD3968	201611B 11/15/2018
0.20	SD3969	201604C 04/05/2018
0.08 DGS	N/A	AG626604 09/22/2018

Flow Calibration Performed By

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete
- Flow Column #
- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #
32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Quality Checks Cont. Performed By DERR

Simulator Temperatures °C

External Digital Therm. ID#: 300918
- ☑ 34°C +-.2 Serial #: SD3967
- ☑ 34°C +-.2 Serial #: SD3968
- ☑ 34°C +-.2 Serial #: SD3969

Calibration Adjustment Performed By

- ☑ Calibration Adjustment N/A
- ☐ Calibration Adjustment Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DERR

- ☑ Barometric Pressure 1017 Gauge
- ID# 68639 1016 Instrument

Mouth Alcohol Solution Lot # 2016A
Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	SD3965
Interferent	SD3966
0.05	SD3967
0.08	SD3968
0.20	SD3969

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Calibration Adjustment
- ☐ Post-Stability Tests
- ☐ Other

Notes/Suggested Service: E-MAILED 6-12-2017

Repeated 0.200 after tightening the external thermostat cover (DERR Seal)

DN/DC DKGJM 6/22/2017

Quality Control Review

APPROVED 6/26/17

Date

- ☑ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☑ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☑ Conduct an Agency Inspection Before Evidentiary Use

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities 2nd Test	80-000874	Miami Police Department	6/12/2017	

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☐	0.077 to 0.083 ☐	0.194 to 0.206 ☑	0.077 to 0.083 ☐
X	X	MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000874 06/12/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:30 Control Test 0.196 06:31 Air Blank 0.000 06:32 Control Test 0.196 06:32 Air Blank 0.000 06:33 Control Test 0.197 06:34 Air Blank 0.000 06:34 Control Test Stats Average 0.1963 Std Dev 0.0006 Rel Std Dev(%) 0.2941 Operator's Signature	X

PG

PGM

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000874	Miami Police Department	6/12/2017	

0.05g/210L

SN: SD3967 Temp: 34.07c

0.047 to 0.053 ☑

MIAMI PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000874
06/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:15
Control Test	0.049	06:16
Air Blank	0.000	06:17
Control Test	0.049	06:17
Air Blank	0.000	06:18
Control Test	0.049	06:19
Air Blank	0.000	06:19

Control Test Stats

Average	0.0490
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DERR
Operator's Signature

0.08g/210L

SN: SD3968 Temp: 34.08c

0.077 to 0.083 ☑

MIAMI PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000874
06/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:20
Control Test	0.079	06:21
Air Blank	0.000	06:22
Control Test	0.080	06:22
Air Blank	0.000	06:23
Control Test	0.080	06:24
Air Blank	0.000	06:24

Control Test Stats

Average	0.0797
Std Dev	0.0006
Rel Std Dev(%)	0.7247

DERR
Operator's Signature

0.20g/210L

SN: SD3969 Temp:34.06c

0.194 to 0.206 ☑

MIAMI PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000874
06/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:25
Control Test	0.193	06:26
Air Blank	0.000	06:26
Control Test	0.197	06:27
Air Blank	0.000	06:28
Control Test	0.196	06:28
Air Blank	0.000	06:29

Control Test Stats

Average	0.1953
Std Dev	0.0021
Rel Std Dev(%)	1.0657

Operator's Signature

DGS 0.08g/210L

Lot AG626604

0.077 to 0.083 ☑

MIAMI PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000874
06/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:35
Control Test	0.079	06:36
Air Blank	0.000	06:36
Control Test	0.078	06:37
Air Blank	0.000	06:37
Control Test	0.078	06:37
Air Blank	0.000	06:38

Control Test Stats

Average	0.0783
Std Dev	0.0006
Rel Std Dev(%)	0.7370

DERR
Operator's Signature