



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FWC - Tampa S/N 80-000901
Date In 6/23/17 Date Out 4/27/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Lab Temp °C 20.0
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 155
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 0.140 (.139 -.169)
 36mm 0.160 (.156 -.190)
 53mm 0.230 (.228 -.278)
 103mm 0.507 (.447 -.547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD3962</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>
0.20	<u>SD3933</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG1621604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 -.169)
 36mm _____ (.156 -.190)
 53mm _____ (.228 -.278)
 103mm _____ (.447 -.547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By [Signature]

Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD3964
 34°C +/- .2 Serial #: SD3933

RECEIVED
JUN 29 2017
FDLE Alcohol Testing Program

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1014 Gauge
 ID# 28427 1013 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>SD3962</u>
0.08	<u>SD3964</u>
0.20	<u>SD3933</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
Sending to CMI for repair to digital dry gas sensor reading 0. Otherwise functioning well. [Signature]
ATP/PC OK SP
[Signature] 6/29/17
 Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Stability Checks # 80-000901 FFWCC Tampa 4/27/17

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
06/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.050	09:37
Air Blank	0.000	09:38
Control Test	0.050	09:39
Air Blank	0.000	09:39
Control Test	0.050	09:40
Air Blank	0.000	09:40
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
06/27/2017
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:42
Control Test	0.080	09:42
Air Blank	0.000	09:43
Control Test	0.079	09:44
Air Blank	0.000	09:44
Control Test	0.079	09:45
Air Blank	0.000	09:45
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	


Operator's Signature

FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
06/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.197	09:47
Air Blank	0.000	09:48
Control Test	0.198	09:48
Air Blank	0.000	09:49
Control Test	0.199	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	


Operator's Signature


FFWCC TAMPA
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000901
06/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.079	09:54
Air Blank	0.000	09:54
Control Test	0.079	09:54
Air Blank	0.000	09:55
Control Test	0.080	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	


Operator's Signature


Operator's Signature