



INSTRUMENT PROCESSING SHEET

Agency Brevard County SO

S/N 80-000940

Florida Department of Law Enforcement

Date In 11/1/2017

DI Completion Date 11/8/17

Ship P/U H/D CMI EE

Intake Performed By <u>SP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>148</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD3962</u></td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>SD1013</u></td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG62660X5</u> 9-22-18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD3962</u>	201707D 07/25/2019	0.080	<u>SD1013</u>	201707E 07/25/2019	0.200	<u>DR3856</u>	201707C 07/24/2019	0.080 DGS	N/A	<u>AG62660X5</u> 9-22-18	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	<u>SD3962</u>	201707D 07/25/2019															
0.080	<u>SD1013</u>	201707E 07/25/2019															
0.200	<u>DR3856</u>	201707C 07/24/2019															
0.080 DGS	N/A	<u>AG62660X5</u> 9-22-18															

Final Release Date

NOV 14 2017

Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.3</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1013</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR3856</u>
---	--

Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.300																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.050																																																
0.080																																																
0.200																																																
0.080 DGS	N/A																																															

Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1016</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>SD1019</u></td> </tr> <tr> <td>Interferent</td> <td><u>SD1021</u></td> </tr> <tr> <td>0.050</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.080</td> <td><u>SD1013</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>SD1019</u>	Interferent	<u>SD1021</u>	0.050	<u>SD3962</u>	0.080	<u>SD1013</u>	0.200	<u>DR3856</u>
Simulator	Serial Number											
0.000	<u>SD1019</u>											
Interferent	<u>SD1021</u>											
0.050	<u>SD3962</u>											
0.080	<u>SD1013</u>											
0.200	<u>DR3856</u>											

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____

Notes/Suggested Service: _____

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date <u>11/8/17</u> <u>JJ DeLeon</u> Admin Review / Date <u>11/14/17</u>

STABILITY CHECKS - # 80-000940

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:14
Control Test	0.050	09:15
Air Blank	0.000	09:15
Control Test	0.050	09:16
Air Blank	0.000	09:16
Control Test	0.050	09:17
Air Blank	0.000	09:18
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

SP
 Operator's Signature

11/14/17
 JD

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.082	09:10
Air Blank	0.000	09:10
Control Test	0.082	09:11
Air Blank	0.000	09:11
Control Test	0.081	09:12
Air Blank	0.000	09:12
Control Test Stats		
Average	0.0817	
Std Dev	0.0005	
Rel. Std Dev(%)	0.7070	

SP
 Operator's Signature

2017-A
 SD1019
 SD1021
 SD3962
 SD1013
 DR385

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	0.199	09:04
Air Blank	0.000	09:04
Control Test	0.198	09:05
Air Blank	0.000	09:05
Control Test	0.199	09:06
Air Blank	0.000	09:06
Control Test	0.199	09:07
Air Blank	0.000	09:07
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2906	

SP
 Operator's Signature

BREUARD COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000940
 11/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:59
Control Test	0.079	08:59
Air Blank	0.000	09:00
Control Test	0.079	09:00
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:01
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7277	

SP
 Operator's Signature

DGS

SP