

INSTRUMENT PROCESSING SHEET

Agency Collier County Sheriff's Office S/N 80-000942

Date In 02/28/2017 Date Out 03/03/2017 Ship P/U H/D CMI EE

<p>Intake Performed By <u>DELL</u></p> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input checked="" type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____	<p>Quality Checks Performed By <u>DELL</u></p> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>227</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32mm <u>156</u> (.139 - .169) 36mm <u>183</u> (.156 - .190) 53mm <u>250</u> (.228 - .278) 103mm <u>511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>66639</u> <input checked="" type="checkbox"/> Stability Checks	<p>Flow Calibration Performed By _____</p> <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 170mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)															
<p>Visual Inspection:</p> <p><u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose</p> <p>Other Equipment:</p> <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>ANTI STATIC BAG</u> <p>Notes: _____</p>	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3967</u></td> <td><u>201603D</u> <u>03/08/2018</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3968</u></td> <td><u>201601F</u> <u>01/26/2018</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3969</u></td> <td><u>201604C</u> <u>04/05/2018</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626605</u> <u>09/03/2018</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	<u>SD3967</u>	<u>201603D</u> <u>03/08/2018</u>	0.08	<u>SD3968</u>	<u>201601F</u> <u>01/26/2018</u>	0.20	<u>SD3969</u>	<u>201604C</u> <u>04/05/2018</u>	0.08 DGS	N/A	<u>AG626605</u> <u>09/03/2018</u>
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD3967</u>	<u>201603D</u> <u>03/08/2018</u>															
0.08	<u>SD3968</u>	<u>201601F</u> <u>01/26/2018</u>															
0.20	<u>SD3969</u>	<u>201604C</u> <u>04/05/2018</u>															
0.08 DGS	N/A	<u>AG626605</u> <u>09/03/2018</u>															

RECEIVED
MAR 08 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

 Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELL

 Barometric Pressure 1024 Gauge
 ID# 28199 1025 Instrument

Mouth Alcohol Solution Lot # 2016A
 Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	<u>SD3965</u>
Interferent	<u>SD3966</u>
0.05	<u>SD3967</u>
0.08	<u>SD3968</u>
0.20	<u>SD3969</u>

Attachments

 Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED**
03/03/2017 **APPROVED**
DA/OC OKAGM 3/8/2017

Brett Kurland

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

3/8/17
Date

Quality Control Review

Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000942	Collier County Sheriff's Office	3/03/2017	<i>PK</i>

PK

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

COLLIER COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-000942
03/03/2017
Software: 8100.27

COLLIER COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-000942
03/03/2017
Software: 8100.27

COLLIER COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-000942
03/03/2017
Software: 8100.27

COLLIER COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-000942
03/03/2017
Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:12
Control Test 0.051 08:13
Air Blank 0.000 08:14
Control Test 0.050 08:14
Air Blank 0.000 08:15
Control Test 0.050 08:16
Air Blank 0.000 08:16

Control Test Stats
Average 0.0503
Std Dev 0.0006
Rel Std Dev(%) 1.1471

Test g/210L Time

Air Blank 0.000 08:18
Control Test 0.081 08:19
Air Blank 0.000 08:19
Control Test 0.081 08:20
Air Blank 0.000 08:20
Control Test 0.082 08:21
Air Blank 0.000 08:22

Control Test Stats
Average 0.0813
Std Dev 0.0006
Rel Std Dev(%) 0.7099

Test g/210L Time

Air Blank 0.000 08:24
Control Test 0.202 08:25
Air Blank 0.000 08:25
Control Test 0.201 08:26
Air Blank 0.000 08:27
Control Test 0.202 08:27
Air Blank 0.000 08:28

Control Test Stats
Average 0.2017
Std Dev 0.0006
Rel Std Dev(%) 0.2863

Test g/210L Time

Air Blank 0.000 08:31
Control Test 0.079 08:31
Air Blank 0.000 08:31
Control Test 0.080 08:32
Air Blank 0.000 08:32
Control Test 0.079 08:33
Air Blank 0.000 08:33

Control Test Stats
Average 0.0793
Std Dev 0.0006
Rel Std Dev(%) 0.7277

ppom

bell
Operator's Signature

bell
Operator's Signature

bell
Operator's Signature

bell
Operator's Signature