



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange County
Date In 4/3/17 Date Out 4/12/17

S/N 30-000964
 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case OK Handle OK
 Dry Gas Holder OK Feet OK
 Keyboard/Plug OK Back/Plugs OK
 Screws tight OK Breath Hose OK

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 199
 Flow Verification (L/s)
 Flow Column # AP105
 32mm 149 (.139 - .169)
 36mm 171 (.156 - .190)
 53mm 258 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR 2035</u>	<u>201605D</u> 3/8/18
0.08	<u>SD 3964</u>	<u>201611B</u> 11/15/18
0.20	<u>SD1025</u>	<u>201604C</u> 4/5/18
0.08 DGS	N/A	<u>AG676609</u> 7/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
APR 12 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure
 ID# 28421 Gauge 1021
1021 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR 2035</u>
0.08	<u>SD 3964</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
Quality Control Review

4/12/17
Date

80-000964
 Stability checks
 4/12/17

INTOXILYZER 8000
 Instrument Initialization
 07:34 04/12/2017

ORANGE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000964
 04/12/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:01
Control Test	0.080	08:01
Air Blank	0.000	08:02
Control Test	0.079	08:02
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DSS

Operator's Signature

ORANGE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000964
 04/12/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:05
Control Test	0.078	08:05
Air Blank	0.000	08:06
Control Test	0.079	08:07
Air Blank	0.000	08:07
Control Test	0.079	08:08
Air Blank	0.000	08:08
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

[Signature]

Operator's Signature

ORANGE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000964
 04/12/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:09
Control Test	0.048	08:10
Air Blank	0.000	08:11
Control Test	0.048	08:11
Air Blank	0.000	08:12
Control Test	0.048	08:12
Air Blank	0.000	08:13
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]

Operator's Signature

ORANGE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000964
 04/12/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:20
Control Test	0.200	08:21
Air Blank	0.000	08:21
Control Test	0.199	08:22
Air Blank	0.000	08:23
Control Test	0.200	08:23
Air Blank	0.000	08:24
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

SP

[Signature]

Operator's Signature

JP