

INSTRUMENT PROCESSING SHEET

RECEIVED

Agency Osceola County S/N 80-000968 AUG 21 2017
 Date In 8/4/17 Date Out 8/21/17 Ship P/U H/D CMI FEE

Intake Performed By DOB

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By DOB

Lab Temp °C 22.5
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 189
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.160 (.139 - .169)
 36mm 0.171 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201603D 3/8/18
0.08	SD1013	201611B 11/5/18
0.20	SD3933	201702B 2/23/19
0.08 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By DOB
Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: SD3933

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DOB

Barometric Pressure 1019 Gauge
 ID# 26932 1018 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD3962
0.08	SD1013
0.20	SD3933

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
Please change menu level 2 to something unique DOB
PA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review _____ Date 8/24/17

Stability Checks # 80-000968 Osceola Corrections 8/21/17 RWS

RWS

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
08/21/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:17
Control Test	0.051	11:17
Air Blank	0.000	11:18
Control Test	0.050	11:19
Air Blank	0.000	11:19
Control Test	0.050	11:20
Air Blank	0.000	11:20
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

RWS

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
08/21/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:22
Control Test	0.083	11:22
Air Blank	0.000	11:23
Control Test	0.081	11:24
Air Blank	0.000	11:24
Control Test	0.082	11:25
Air Blank	0.000	11:25
Control Test Stats		
Average	0.0820	
Std Dev	0.0010	
Rel Std Dev(%)	1.2195	

RWS

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
08/21/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:27
Control Test	0.205	11:27
Air Blank	0.000	11:28
Control Test	0.204	11:28
Air Blank	0.000	11:29
Control Test	0.204	11:30
Air Blank	0.000	11:30
Control Test Stats		
Average	0.2043	
Std Dev	0.0006	
Rel Std Dev(%)	0.2826	

RWS

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
08/21/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:31
Control Test	0.079	11:32
Air Blank	0.000	11:32
Control Test	0.080	11:33
Air Blank	0.000	11:33
Control Test	0.079	11:33
Air Blank	0.000	11:34
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

RWS

Operator's Signature

SP

RWS
8/21/17