

INSTRUMENT PROCESSING SHEET

Agency Pinellas County S/N 80-001005
 Date In 1/18/17 Date Out 1/18/17 Ship P/U H/D CMI EE

| Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____ | Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>221</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.152</u> (.139 - .169) 36mm <u>0.167</u> (.156 - .190) 53mm <u>0.238</u> (.228 - .278) 103mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3962</u></td> <td><u>201603D</u> <u>03/08/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3933</u> ⁶⁴</td> <td><u>201601F</u> <u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG2605</u> <u>9/22/18</u></td> </tr> </tbody> </table> | Simulator | Serial # | Lot #/Exp | 0.05 | <u>SD3962</u> | <u>201603D</u> <u>03/08/18</u> | 0.08 | <u>SD3933</u> ⁶⁴ | <u>201601F</u> <u>1/26/18</u> | 0.20 | <u>SD3933</u> | <u>201604C</u> <u>4/5/18</u> | 0.08 DGS | N/A | <u>AG2605</u> <u>9/22/18</u> | Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------|----------|-----------|------|---------------|-----------------------------------|------|--------------------------------|----------------------------------|------|---------------|---------------------------------|----------|-----|---------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Simulator | Serial # | Lot #/Exp | | | | | | | | | | | | | | | |
| 0.05 | <u>SD3962</u> | <u>201603D</u> <u>03/08/18</u> | | | | | | | | | | | | | | | |
| 0.08 | <u>SD3933</u> ⁶⁴ | <u>201601F</u> <u>1/26/18</u> | | | | | | | | | | | | | | | |
| 0.20 | <u>SD3933</u> | <u>201604C</u> <u>4/5/18</u> | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | <u>AG2605</u> <u>9/22/18</u> | | | | | | | | | | | | | | | |
| Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ | | | | | | | | | | | | | | | | | |
| Suggested Service _____ _____ | | | | | | | | | | | | | | | | | |

RECEIVED
 JAN 19 2017
 FDLE
 Alcohol Testing Program

| Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> | Simulator | Serial Number | Lot Number | Expiration | 0.000 | | N/A | N/A | 0.040 | | | | 0.100 | | | | 0.200 | | | | 0.400 | | | | 0.080 DGS | N/A | | | Simulator | Serial Number | Lot Number | Expiration | 0.05 | | | | 0.08 | | | | 0.20 | | | | 0.08 DGS | N/A | | | Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure _____ Gauge ID# <u>26932</u> _____ ID# _____ Instrument Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2834</u></td> </tr> <tr> <td>0.05</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> </tr> </tbody> </table> | Simulator | Serial Number | 0.00 | <u>G2880</u> | Interferent | <u>G2834</u> | 0.05 | <u>SD3962</u> | 0.08 | <u>SD3964</u> | 0.20 | <u>SD3933</u> |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------|---------------|------------|------------|-------|--|-----|-----|-------|--|--|--|-------|--|--|--|-------|--|--|--|-------|--|--|--|-----------|-----|--|--|-----------|---------------|------------|------------|------|--|--|--|------|--|--|--|------|--|--|--|----------|-----|--|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------|---------------|------|--------------|-------------|--------------|------|---------------|------|---------------|------|---------------|
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.000 | | N/A | N/A | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.040 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.100 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.200 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.400 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.080 DGS | N/A | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.05 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.20 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Simulator | Serial Number | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.00 | <u>G2880</u> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Interferent | <u>G2834</u> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.05 | <u>SD3962</u> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 | <u>SD3964</u> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.20 | <u>SD3933</u> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____ | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

Notes: QA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hankland
 Quality Control Review

1/19/17
 Date

Stability Checks # 80-001005 Pinellas County S.O. 1/18/17 *DBS*

DBS

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001005
01/18/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:48 |
| Control Test | 0.048 | 13:48 |
| Air Blank | 0.000 | 13:49 |
| Control Test | 0.049 | 13:50 |
| Air Blank | 0.000 | 13:50 |
| Control Test | 0.050 | 13:51 |
| Air Blank | 0.000 | 13:51 |
| Control Test Stats | | |
| Average | 0.0490 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 2.0408 | |

DBS
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001005
01/18/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:54 |
| Control Test | 0.079 | 13:55 |
| Air Blank | 0.000 | 13:55 |
| Control Test | 0.080 | 13:56 |
| Air Blank | 0.000 | 13:56 |
| Control Test | 0.079 | 13:57 |
| Air Blank | 0.000 | 13:58 |
| Control Test Stats | | |
| Average | 0.0793 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7277 | |

DBS
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001005
01/18/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:59 |
| Control Test | 0.198 | 13:59 |
| Air Blank | 0.000 | 14:00 |
| Control Test | 0.199 | 14:01 |
| Air Blank | 0.000 | 14:01 |
| Control Test | 0.199 | 14:02 |
| Air Blank | 0.000 | 14:02 |
| Control Test Stats | | |
| Average | 0.1987 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2906 | |

DBS
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001005
01/18/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 14:04 |
| Control Test | 0.079 | 14:05 |
| Air Blank | 0.000 | 14:05 |
| Control Test | 0.078 | 14:05 |
| Air Blank | 0.000 | 14:06 |
| Control Test | 0.078 | 14:06 |
| Air Blank | 0.000 | 14:07 |
| Control Test Stats | | |
| Average | 0.0783 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7370 | |

SP BK
DBS
Operator's Signature