



# Stability Checks # 80-001007 Haines City P.D. 8/3/17

DCS

HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001007  
08/03/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:23
Control Test	0.050	13:24
Air Blank	0.000	13:24
Control Test	0.049	13:25
Air Blank	0.000	13:26
Control Test	0.049	13:26
Air Blank	0.000	13:27
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1703	

*[Signature]*  
Operator's Signature

8/4/17

HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001007  
08/03/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:18
Control Test	0.078	13:18
Air Blank	0.000	13:19
Control Test	0.079	13:20
Air Blank	0.000	13:20
Control Test	0.080	13:21
Air Blank	0.000	13:22
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel. Std Dev(%)	1.2658	

*[Signature]*

*[Signature]*  
Operator's Signature

HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001007  
08/03/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:12
Control Test	0.200	13:13
Air Blank	0.000	13:13
Control Test	0.200	13:14
Air Blank	0.000	13:15
Control Test	0.200	13:15
Air Blank	0.000	13:16
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

*[Signature]*  
Operator's Signature

HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001007  
08/03/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:29
Control Test	0.081	13:30
Air Blank	0.000	13:30
Control Test	0.082	13:31
Air Blank	0.000	13:31
Control Test	0.081	13:31
Air Blank	0.000	13:32
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7099	

*[Signature]*  
Operator's Signature