



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Palm Bay PD S/N 80-001009
Date In 7/17/17 Date Out 7/20/17 Ship P/U H/D COME EE

RECEIVED
ALCOHOL TESTING PROGRAM
FDLE
20 2017

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Lab Temp °C 23.9
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 187
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm .152 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .240 (.228 - .278)
 103mm .519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421
 Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD3962</u>	<u>201603D</u> <u>3-8-18</u>
0.08	<u>SD1013</u>	<u>201611B</u> <u>11-15-18</u>
0.20	<u>SD3933</u>	<u>201702B</u> <u>2-23-19</u>
0.08 DGS	N/A	<u>AG62660+</u> <u>9-22-18</u>

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By SP
Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: SD3933

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By ST

Barometric Pressure 1016 Gauge
 ID# 28427 1014 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>SD3962</u>
0.08	<u>SD1013</u>
0.20	<u>SD3933</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
QA/QC v 8015 7/20/17

[Signature]
 Quality Control Review

7/20/17
 Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

STABILITY CHECKS - ~~LEE COUNTY~~ 50 - 80 - 001009
 Palm Bay P.D. SP
 2/10/18

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.050	09:06
Air Blank	0.000	09:06
Control Test	0.049	09:07
Air Blank	0.000	09:08
Control Test	0.050	09:08
Air Blank	0.000	09:09
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:59
Control Test	0.079	09:00
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:02
Control Test	0.079	09:03
Air Blank	0.000	09:03
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.199	08:52
Air Blank	0.000	08:53
Control Test	0.199	08:53
Air Blank	0.000	08:54
Control Test	0.199	08:55
Air Blank	0.000	08:55
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001009
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.079	08:49
Air Blank	0.000	08:49
Control Test	0.078	08:49
Air Blank	0.000	08:50
Control Test	0.079	08:50
Air Blank	0.000	08:51
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DAS

SP

Operator's Signature

SP

STABILITY CHECKS - LEE COUNTY SO - 80-001009

PALM BAY P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001009
07/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.050	09:06
Air Blank	0.000	09:06
Control Test	0.049	09:07
Air Blank	0.000	09:08
Control Test	0.050	09:08
Air Blank	0.000	09:09
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1625	

SP

Operator's Signature

PALM BAY P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001009
07/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:59
Control Test	0.079	09:00
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:02
Control Test	0.079	09:03
Air Blank	0.000	09:03
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7277	

SP

Operator's Signature

PALM BAY P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001009
07/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.199	08:52
Air Blank	0.000	08:53
Control Test	0.199	08:53
Air Blank	0.000	08:54
Control Test	0.199	08:55
Air Blank	0.000	08:55
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

SP

Operator's Signature

PALM BAY P.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001009
07/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.079	08:49
Air Blank	0.000	08:49
Control Test	0.078	08:49
Air Blank	0.000	08:50
Control Test	0.079	08:50
Air Blank	0.000	08:51
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7339	

DGS

SP

Operator's Signature

SP