



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Clay County SO

S/N 80-001071

Date In 1/4/17

Date Out 2/15/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 165
 Flow Verification (L/s)
 Flow Column # A7P103
 32mm 146 (.139 - .169)
 36mm 165 (.156 - .190)
 53mm 240 (.228 - .278)
 103mm 498 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G11739</u> <u>DE2035</u>	<u>201603D</u> <u>318/18</u>
0.08	<u>G8149</u> <u>SD1011</u>	<u>201604F</u> <u>1/24/18</u>
0.20	<u>SD1025</u> <u>G11621</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG-626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By [Signature]

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 15 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1004 Gauge
 ID# 26932 1003 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B 2017-A

Simulator	Serial Number
0.00	<u>SD1011 G2879</u>
Interferent	<u>SD1021 G8144</u>
0.05	<u>DE2035 G11739</u>
0.08	<u>SD1011 G8149</u>
0.20	<u>SD1025 G11621</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: Change level 2 Pw to unique P.W.
QC

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

2/15/17
Date

80-001071
Stability Checks
2/15/17

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
02/15/2017
Software: 8100.27

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:23
Control Test	0.079	08:24
Air Blank	0.000	08:25
Control Test	0.079	08:25
Air Blank	0.000	08:26
Control Test	0.079	08:26
Air Blank	0.000	08:27
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Test	g/210L	Time
Air Blank	0.000	08:15
Control Test	0.078	08:16
Air Blank	0.000	08:16
Control Test	0.079	08:17
Air Blank	0.000	08:17
Control Test	0.079	08:17
Air Blank	0.000	08:18
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:19
Control Test	0.049	08:20
Air Blank	0.000	08:20
Control Test	0.049	08:21
Air Blank	0.000	08:21
Control Test	0.049	08:22
Air Blank	0.000	08:22
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature 

Operator's Signature 

Operator's Signature 

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:28
Control Test	0.196	08:29
Air Blank	0.000	08:29
Control Test	0.198	08:30
Air Blank	0.000	08:30
Control Test	0.199	08:31
Air Blank	0.000	08:32
Control Test Stats		
Average	0.1977	
Std Dev	0.0015	
Rel Std Dev(%)	0.7728	

DGS
YBK

Operator's Signature 