

INSTRUMENT PROCESSING SHEET

Agency FHP - Deland S/N 80-001119
 Date In 3/13/17 Date Out 3/23/2017 Ship P/U H/D CMI EE

Intake Performed By 108

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By 909M

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 203
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 1156 (.139 - .169)
 36mm 1167 (.156 - .190)
 53mm 224 (.228 - .278)
 103mm 497 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/5/18</u>
0.08	<u>SD3964</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>RR26605</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAR 23 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By 909M

Barometric Pressure
 ID# 28421 Gauge 1022
 Instrument 1024

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Huckland

Quality Control Review

3/23/17
 Date

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001119
 03/23/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.051	09:32
Air Blank	0.000	09:33
Control Test	0.052	09:34
Air Blank	0.000	09:34
Control Test	0.051	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0513	
Std Dev	0.0006	
Rel Std Dev(%)	1.1247	

P. Murphy
 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001119
 03/23/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:27
Control Test	0.081	09:27
Air Blank	0.000	09:28
Control Test	0.082	09:29
Air Blank	0.000	09:29
Control Test	0.082	09:30
Air Blank	0.000	09:30
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

P. Murphy
 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001119
 03/23/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:21
Control Test	0.201	09:21
Air Blank	0.000	09:22
Control Test	0.200	09:23
Air Blank	0.000	09:23
Control Test	0.200	09:24
Air Blank	0.000	09:24
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

P. Murphy
 Operator's Signature

SP
 BK

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001119
 03/23/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.079	09:38
Air Blank	0.000	09:38
Control Test	0.080	09:39
Air Blank	0.000	09:39
Control Test	0.079	09:39
Air Blank	0.000	09:40
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

P. Murphy
 Operator's Signature