

80-001128

State. lity Clecks

11/2/17

INTOXILYZER 8000
Instrument Initialization
06:55 11/02/2017

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001128
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.203	09:05
Air Blank	0.000	09:06
Control Test	0.202	09:07
Air Blank	0.000	09:07
Control Test	0.202	09:08
Air Blank	0.000	09:08
Control Test Stats		
Average	0.2023	
Std Dev	0.0006	
Rel Std Dev(%)	0.2853	

Operator's Signature

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001128
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:01
Control Test	0.081	09:02
Air Blank	0.000	09:02
Control Test	0.080	09:03
Air Blank	0.000	09:03
Control Test	0.080	09:03
Air Blank	0.000	09:04
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

Operator's Signature

11/2/17

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001128
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.080	09:10
Air Blank	0.000	09:11
Control Test	0.081	09:11
Air Blank	0.000	09:12
Control Test	0.081	09:13
Air Blank	0.000	09:13
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Operator's Signature

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001128
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.051	09:15
Air Blank	0.000	09:16
Control Test	0.051	09:16
Air Blank	0.000	09:17
Control Test	0.051	09:18
Air Blank	0.000	09:18
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature