

INSTRUMENT PROCESSING SHEET

Agency Orange City PD S/N 80-001159
 Date In 8/3/17 Date Out 9/1/2017 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case _____ Handle _____ <u>ok</u> Dry Gas Holder _____ <u>ok</u> Feet _____ <u>ok</u> Keyboard/Plug _____ <u>ok</u> Back/Plugs _____ <u>ok</u> Screws tight _____ <u>ok</u> Breath Hose _____ Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: <u>Right side of handle not attached to instrument.</u>	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.1</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>226</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.160</u> (.139 - .169) 36mm <u>.175</u> (.156 - .190) 53mm <u>.246</u> (.228 - .278) 103mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 57mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)															
	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3962</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1013</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG69605</u> <u>7/14/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD3962</u>	<u>201707D</u> <u>7/25/19</u>	0.08	<u>SD1013</u>	<u>201707E</u> <u>7/25/19</u>	0.20	<u>SD3933</u>	<u>201707C</u> <u>7/24/19</u>	0.08 DGS	N/A	<u>AG69605</u> <u>7/14/18</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By <u>[Signature]</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1013</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3933</u>
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD3962</u>	<u>201707D</u> <u>7/25/19</u>															
0.08	<u>SD1013</u>	<u>201707E</u> <u>7/25/19</u>															
0.20	<u>SD3933</u>	<u>201707C</u> <u>7/24/19</u>															
0.08 DGS	N/A	<u>AG69605</u> <u>7/14/18</u>															

Calibration Adjustment Performed By _____ <input checked="" type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____			
Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks			
Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1014</u> Gauge ID# <u>26932</u> <u>1013</u> Instrument	
Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u>	
Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>SD3962</u>
0.08	<u>SD1013</u>
0.20	<u>SD3933</u>
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>form 40</u>	

Notes/Suggested Service:
DC [Signature] 9/13/17
[Signature]
[Signature]
 Quality Control Review [Signature] Date 9/13/17

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
08/31/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:32
Control Test	0.050	13:32
Air Blank	0.000	13:33
Control Test	0.050	13:33
Air Blank	0.000	13:34
Control Test	0.049	13:35
Air Blank	0.000	13:35
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

P Murphy
Operator's Signature

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
08/31/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:26
Control Test	0.081	13:27
Air Blank	0.000	13:28
Control Test	0.082	13:28
Air Blank	0.000	13:29
Control Test	0.080	13:29
Air Blank	0.000	13:30
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

P Murphy
Operator's Signature

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
08/31/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:21
Control Test	0.196	13:22
Air Blank	0.000	13:23
Control Test	0.198	13:23
Air Blank	0.000	13:24
Control Test	0.200	13:24
Air Blank	0.000	13:25
Control Test Stats		
Average	0.1980	
Std Dev	0.0020	
Rel Std Dev(%)	1.0101	

P Murphy
Operator's Signature

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
08/31/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:58
Control Test	0.082	13:58
Air Blank	0.000	13:58
Control Test	0.083	13:59
Air Blank	0.000	13:59
Control Test	0.083	13:59
Air Blank	0.000	14:00
Control Test Stats		
Average	0.0827	
Std Dev	0.0006	
Rel Std Dev(%)	0.6984	

DGS

P Murphy
Operator's Signature

DGS

9/3/17
JD

INSTRUMENT PROCESSING SHEET

Agency Orange City S/N 80-001159
 Date In 2/24/17 Date Out 2/28/17 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>220</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.160</u> (.139 - .169) 36mm <u>.171</u> (.156 - .190) 53mm <u>.238</u> (.228 - .278) 103mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G11739</u></td> <td><u>201603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>G8149</u></td> <td><u>201601F</u> <u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>G11621</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626605</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>G11739</u>	<u>201603D</u> <u>3/8/18</u>	0.08	<u>G8149</u>	<u>201601F</u> <u>1/26/18</u>	0.20	<u>G11621</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	N/A	<u>AG626605</u> <u>9/22/18</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - <u>7mm</u> <input type="checkbox"/> 15L/min - <u>53mm</u> <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>G11739</u>	<u>201603D</u> <u>3/8/18</u>															
0.08	<u>G8149</u>	<u>201601F</u> <u>1/26/18</u>															
0.20	<u>G11621</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	N/A	<u>AG626605</u> <u>9/22/18</u>															
Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____																	

RECEIVED
 MAR 14 2017
 FDLE Alcohol Testing Program

Optical Bench Calibration Performed By <u>[Signature]</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1024</u> ID # <u>28427</u>																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000	<u>SD1016</u>	N/A	N/A																				
0.040	<u>SD1024</u>	<u>16101</u>	<u>2/2/18</u>																				
0.100	<u>SD1022</u>	<u>16001</u>	<u>5/8/18</u>																				
0.200	<u>DR3855</u>	<u>16103</u>	<u>6/14/18</u>																				
0.400	<u>G2407</u>	<u>16102</u>	<u>3/22/18</u>																				
0.080 DGS	N/A	<u>15615080A2</u>	<u>7/5/17</u>																				
<input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G11739</u></td> <td><u>201603D</u></td> <td><u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>G8149</u></td> <td><u>201601F</u></td> <td><u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>G11621</u></td> <td><u>201604C</u></td> <td><u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626605</u></td> <td><u>9/22/18</u></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05	<u>G11739</u>	<u>201603D</u>	<u>3/8/18</u>	0.08	<u>G8149</u>	<u>201601F</u>	<u>1/26/18</u>	0.20	<u>G11621</u>	<u>201604C</u>	<u>4/5/18</u>	0.08 DGS	N/A	<u>AG626605</u>	<u>9/22/18</u>
Simulator	Serial Number	Lot Number	Expiration																				
0.05	<u>G11739</u>	<u>201603D</u>	<u>3/8/18</u>																				
0.08	<u>G8149</u>	<u>201601F</u>	<u>1/26/18</u>																				
0.20	<u>G11621</u>	<u>201604C</u>	<u>4/5/18</u>																				
0.08 DGS	N/A	<u>AG626605</u>	<u>9/22/18</u>																				

Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1024</u> Gauge ID# <u>26932</u> <u>1024</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	<u>G2879</u>
Interferent	<u>G8144</u>
0.05	<u>G11739</u>
0.08	<u>G8149</u>
0.20	<u>G11621</u>

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>form 40</u>	
--	--

Notes: QC + DGS

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Butt Kirkland
 Quality Control Review

3/14/17
 Date

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
02/28/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:28	
Control Test	0.051	09:28	
Air Blank	0.000	09:29	
Control Test	0.049	09:30	
Air Blank	0.000	09:30	
Control Test	0.050	09:31	
Air Blank	0.000	09:31	
Control Test Stats			
Average	0.0500		
Std Dev	0.0010		
Rel Std Dev(%)	2.0000		

P Murphy
Operator's Signature

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
02/28/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:23	
Control Test	0.082	09:24	
Air Blank	0.000	09:24	
Control Test	0.081	09:25	
Air Blank	0.000	09:25	
Control Test	0.082	09:26	
Air Blank	0.000	09:27	
Control Test Stats			
Average	0.0817		
Std Dev	0.0006		
Rel Std Dev(%)	0.7070		

P Murphy
Operator's Signature

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
02/28/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:37	
Control Test	0.201	09:38	
Air Blank	0.000	09:39	
Control Test	0.201	09:39	
Air Blank	0.000	09:40	
Control Test	0.200	09:40	
Air Blank	0.000	09:41	
Control Test Stats			
Average	0.2007		
Std Dev	0.0006		
Rel Std Dev(%)	0.2877		

P Murphy
Operator's Signature

DGS
T-X

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
02/28/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:33	
Control Test	0.078	09:33	
Air Blank	0.000	09:34	
Control Test	0.078	09:34	
Air Blank	0.000	09:35	
Control Test	0.079	09:35	
Air Blank	0.000	09:36	
Control Test Stats			
Average	0.0783		
Std Dev	0.0006		
Rel Std Dev(%)	0.7370		

DGS

P Murphy
Operator's Signature

ORANGE CITY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001159
03/07/2017 10:03:28

Auto Calibration
Max Power Res Value = 28
Auto Range Res Value = 16

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12714, Sum Io = 14233

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1100 (-0.0160)
Sample #2 = 0.1040 (0.0370)
Sample #3 = 0.0970 (0.0440)
Sample #4 = 0.1060 (0.0510)
Avg % Abs = 0.1023 (0.0440)
STD DEV = 0.0047 (0.0070)
REL STD DEV = 4.618 (15.909)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0960 (-0.0100)
Sample #2 = 0.1090 (0.0030)
Sample #3 = 0.0870 (-0.0060)
Sample #4 = 0.1120 (0.0000)
Avg % Abs = 0.1027 (-0.0010)
STD DEV = 0.0137 (0.0046)
REL STD DEV = 13.296 (458.258)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12701, Sum Io = 14233

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8020 (-0.0390)
Sample #2 = 0.7800 (-0.0090)
Sample #3 = 0.8020 (0.0100)
Sample #4 = 0.7860 (0.0320)
Avg % Abs = 0.7893 (0.0110)
STD DEV = 0.0114 (0.0205)
REL STD DEV = 1.441 (186.530)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5010 (0.0090)
Sample #2 = 1.5160 (0.0080)
Sample #3 = 1.5000 (0.0380)
Sample #4 = 1.5030 (0.0420)
Avg % Abs = 1.5063 (0.0293)
STD DEV = 0.0085 (0.0186)
REL STD DEV = 0.565 (63.352)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12692, Sum Io = 14223

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8130 (-0.0250)
Sample #2 = 1.8590 (-0.0320)
Sample #3 = 1.8550 (-0.0280)
Sample #4 = 1.8140 (0.0030)
Avg % Abs = 1.8427 (-0.0190)
STD DEV = 0.0249 (0.0192)
REL STD DEV = 1.352 (100.828)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5730 (-0.0190)
Sample #2 = 3.5720 (-0.0150)
Sample #3 = 3.6020 (-0.0310)
Sample #4 = 3.5660 (0.0000)
Avg % Abs = 3.5800 (-0.0153)
STD DEV = 0.0193 (0.0155)
REL STD DEV = 0.539 (101.104)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12689, Sum Io = 14222

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4680 (-0.0090)
Sample #2 = 3.4990 (0.0000)
Sample #3 = 3.5160 (-0.0100)
Sample #4 = 3.5330 (-0.0140)
Avg % Abs = 3.5160 (-0.0080)
STD DEV = 0.0170 (0.0072)
REL STD DEV = 0.484 (90.139)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)
Sample #1 = 6.8490 (0.0060)
Sample #2 = 6.9100 (0.0050)
Sample #3 = 6.9090 (0.0090)
Sample #4 = 6.9140 (0.0110)
Avg % Abs = 6.9110 (0.0083)
STD DEV = 0.0026 (0.0031)
REL STD DEV = 0.038 (36.661)

Sol Value = 0.400 g/210L ***
Fit value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12686, Sum Io = 14221

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.7140 (-0.0060)
Sample #2 = 6.6390 (0.0630)
Sample #3 = 6.6570 (0.0740)
Sample #4 = 6.6670 (0.0770)
Avg % Abs = 6.6543 (0.0713)
STD DEV = 0.0142 (0.0074)
REL STD DEV = 0.213 (10.333)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)
Sample #1 = 12.9620 (0.0010)
Sample #2 = 12.9400 (0.0260)
Sample #3 = 12.9610 (0.0370)
Sample #4 = 12.9640 (0.0410)
Avg % Abs = 12.9550 (0.0347)
STD DEV = 0.0131 (0.0078)
REL STD DEV = 0.101 (22.406)

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.102
Std Dev = 0.00 Rel Std Dev = 4.62
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.789
Std Dev = 0.01 Rel Std Dev = 1.44
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.843
Std Dev = 0.02 Rel Std Dev = 1.35
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.516
Std Dev = 0.02 Rel Std Dev = 0.48
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.654
Std Dev = 0.01 Rel Std Dev = 0.21
Zero Order Coef = -247.01
First Order Coef = 2651.15
Second Order Coef = 37.26
Standard Deviation = 22.599211

RWB
13K

80-001159

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 ng/l or 0.000 g/210L
 % Abs = 0.103
 Std Dev = 0.01 Rel Std Dev = 13.30
 Sol Val = 0.1905 ng/l or 0.040 g/210L
 % Abs = 1.506
 Std Dev = 0.01 Rel Std Dev = 0.56
 Sol Val = 0.4762 ng/l or 0.100 g/210L
 % Abs = 3.580
 Std Dev = 0.02 Rel Std Dev = 0.54
 Sol Val = 0.9524 ng/l or 0.200 g/210L
 % Abs = 6.911
 Std Dev = 0.00 Rel Std Dev = 0.04
 Sol Val = 1.9048 ng/l or 0.400 g/210L
 % Abs = 12.955
 Std Dev = 0.01 Rel Std Dev = 0.10
 Zero Order Coef = -111.93
 First Order Coef = 1306.60
 Second Order Coef = 13.26
 Standard Deviation = 24.152010

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0005
0.040	0.039	0.0008
0.100	0.100	-0.0001
0.200	0.200	-0.0002
0.400	0.400	0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0005
0.040	0.040	0.0004
0.100	0.099	0.0006
0.200	0.201	-0.0006
0.400	0.400	0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3021.00
 Sample #2 = 3080.00
 Sample #3 = 3093.00
 Sample #4 = 3018.00
 Average Result = 3063.6667
 STD DEV = 40.0791
 REL STD DEV = 1.308

 ***** CHANNEL 2
 Sample #1 = 3445.00
 Sample #2 = 3440.00
 Sample #3 = 3455.00
 Sample #4 = 3415.00
 Average Result = 3436.6667
 STD DEV = 20.2073
 REL STD DEV = 0.588

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1024
 3 um H2O Adjust (mg/l*10,000) = 746
 9 um H2O Adjust (mg/l*10,000) = 373
 **** AUTO CAL PASS

HSB
 ASK

ORANGE CITY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001159
 03/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:45
Control Test	0.049	10:46
Air Blank	0.000	10:47
Control Test	0.050	10:47
Air Blank	0.000	10:48
Control Test	0.050	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

P. Murphy
 Operator's Signature

ORANGE CITY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001159
 03/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.080	10:51
Air Blank	0.000	10:51
Control Test	0.079	10:52
Air Blank	0.000	10:53
Control Test	0.080	10:53
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

P. Murphy
 Operator's Signature

ORANGE CITY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001159
 03/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:41
Control Test	0.195	10:42
Air Blank	0.000	10:42
Control Test	0.196	10:43
Air Blank	0.000	10:44
Control Test	0.197	10:44
Air Blank	0.000	10:45
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	

P. Murphy
 Operator's Signature

ORANGE CITY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001159
 03/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:55
Control Test	0.080	10:56
Air Blank	0.000	10:56
Control Test	0.080	10:56
Air Blank	0.000	10:57
Control Test	0.080	10:57
Air Blank	0.000	10:58
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

065

P. Murphy
 Operator's Signature

POST CALIBRATION - ADJUST
 STABILITIES

JK

ADP