



INSTRUMENT PROCESSING SHEET

Agency Panama City PDS/N 80-001204

Florida Department of Law Enforcement

Date In 11/27/17DI Completion Date 12/1/17 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>247</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.195</u> (.139 - .169) 36 mm <u>.203</u> (.156 - .190) 53 mm <u>.273</u> (.228 - .278) 103 mm <u>.531</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By <u>[Signature]</u> Flow Column # <u>ATP103</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>247</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.520</u> (.447 - .547)																																																											
Final Release Date <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">DEC 05 2017</div> <div style="text-align: center; font-weight: bold;">Alcohol Testing Program</div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG715202</u> <u>6/1/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>G2835</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.9</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1025</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>G2835</u>	<u>201707D</u> <u>7/25/19</u>																																																											
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>																																																											
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>																																																											
0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>[Signature]</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2840</u></td> </tr> <tr> <td>0.050</td> <td><u>G2835</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2880</u>	Interferent	<u>G2840</u>	0.050	<u>G2835</u>	0.080	<u>DR1279</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	<u>G2880</u>																																																												
Interferent	<u>G2840</u>																																																												
0.050	<u>G2835</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>SD1025</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input type="checkbox"/> Post-Stability Checks																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<div style="text-align: center; font-size: 1.2em;"> <u>SP 12/4/17</u> <u>[Signature]</u> <u>12/5/17</u> Tech Review / Date Admin Review / Date </div>																																																											

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:10
Control Test	0.048	09:11
Air Blank	0.000	09:12
Control Test	0.048	09:12
Air Blank	0.000	09:13
Control Test	0.048	09:13
Air Blank	0.000	09:14
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:04
Control Test	0.079	09:05
Air Blank	0.000	09:05
Control Test	0.079	09:06
Air Blank	0.000	09:06
Control Test	0.079	09:07
Air Blank	0.000	09:08
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:16
Control Test	0.197	09:17
Air Blank	0.000	09:17
Control Test	0.197	09:18
Air Blank	0.000	09:18
Control Test	0.198	09:19
Air Blank	0.000	09:19
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

P Murphy

Operator's Signature

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.080	09:30
Air Blank	0.000	09:31
Control Test	0.080	09:31
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P Murphy

Operator's Signature

PANAMA CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001204
11/30/2017
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff)) = 7.211
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.574
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.758
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 708
Rounded Intercept = -722673
Correlation = 0.99616

P Murphy

SP

12/5/17
DE