

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001249	Broward County Sheriff's Office	4/18/2017	<i>Full</i>

kt

0.05g/210L 0.047 to 0.053 ☒	0.08g/210L 0.077 to 0.083 ☒	0.20g/210L 0.194 to 0.206 ☒	DGS 0.08g/210L 0.077 to 0.083 ☒
BROWARD COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001249 04/18/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:03 Control Test 0.050 06:04 Air Blank 0.000 06:04 Control Test 0.051 06:05 Air Blank 0.000 06:06 Control Test 0.050 06:06 Air Blank 0.000 06:07 Control Test Stats Average 0.0503 Std Dev 0.0006 Rel Std Dev(%) 1.1471	BROWARD COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001249 04/18/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:08 Control Test 0.081 06:08 Air Blank 0.000 06:09 Control Test 0.081 06:10 Air Blank 0.000 06:10 Control Test 0.081 06:11 Air Blank 0.000 06:11 Control Test Stats Average 0.0810 Std Dev 0.0000 Rel Std Dev(%) 0.0000	BROWARD COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001249 04/18/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:12 Control Test 0.200 06:13 Air Blank 0.000 06:14 Control Test 0.198 06:14 Air Blank 0.000 06:15 Control Test 0.197 06:16 Air Blank 0.000 06:16 Control Test Stats Average 0.1983 Std Dev 0.0015 Rel Std Dev(%) 0.7702	BROWARD COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001249 04/18/2017 Software: 8100.27 Test g/210L Time Air Blank 0.000 06:18 Control Test 0.080 06:18 Air Blank 0.000 06:19 Control Test 0.079 06:19 Air Blank 0.000 06:20 Control Test 0.080 06:20 Air Blank 0.000 06:20 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247
Operator's Signature <i>Full</i>	Operator's Signature <i>Full</i>	Operator's Signature <i>Full</i>	Operator's Signature <i>Full</i>

gpm

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 60-001249
04/18/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:44
Control Test	0.081	10:44
Air Blank	0.000	10:44
Control Test	0.080	10:45
Air Blank	0.000	10:45
Control Test	0.079	10:45
Air Blank	0.000	10:46
Control Test	0.080	10:46
Air Blank	0.000	10:47
Control Test	0.080	10:47
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:49
Control Test	0.081	10:50
Air Blank	0.000	10:50
Control Test	0.081	10:51
Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:52
Control Test	0.079	10:52
Air Blank	0.000	10:53
Control Test	0.080	10:53
Air Blank	0.000	10:54
Control Test	0.081	10:54
Air Blank	0.000	10:54
Control Test	0.081	10:55
Air Blank	0.000	10:55
Control Test	0.081	10:56
Air Blank	0.000	10:56
Control Test	0.080	10:56
Air Blank	0.000	10:57
Control Test	0.081	10:57
Air Blank	0.000	10:58
Control Test	0.081	10:58
Air Blank	0.000	10:59
Control Test	0.081	10:59
Air Blank	0.000	10:59
Control Test	0.080	11:00
Air Blank	0.000	11:00
Control Test Stats		
Average	0.0814	
Std Dev	0.0007	
Rel Std Dev(%)	0.8349	

Deer
Operator's Signature

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 60-001249
04/18/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:45
Control Test	0.082	13:45
Air Blank	0.000	13:46
Control Test	0.081	13:46
Air Blank	0.000	13:46
Control Test	0.081	13:47
Air Blank	0.000	13:47
Control Test	0.082	13:48
Air Blank	0.000	13:48
Control Test	0.082	13:48
Air Blank	0.000	13:49
Control Test	0.082	13:49
Air Blank	0.000	13:50
Control Test	0.082	13:50
Air Blank	0.000	13:51
Control Test	0.082	13:51
Air Blank	0.000	13:51
Control Test	0.082	13:52
Air Blank	0.000	13:52
Control Test	0.082	13:53
Air Blank	0.000	13:53
Control Test	0.082	13:53
Air Blank	0.000	13:54
Control Test	0.082	13:54
Air Blank	0.000	13:55
Control Test	0.082	13:55
Air Blank	0.000	13:56
Control Test	0.081	13:56
Air Blank	0.000	13:56
Control Test	0.082	13:57
Air Blank	0.000	13:57
Control Test	0.082	13:57
Air Blank	0.000	13:58
Control Test	0.081	13:58
Air Blank	0.000	13:59
Control Test	0.081	13:59
Air Blank	0.000	14:00
Control Test	0.081	14:00
Air Blank	0.000	14:00
Control Test	0.081	14:01
Air Blank	0.000	14:01
Control Test Stats		
Average	0.0817	
Std Dev	0.0005	
Rel Std Dev(%)	0.5993	

Deer
Operator's Signature

QJm

BK