



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange County S/N 80-001256

Date In 8/22/17 Date Out 8/31/17 Ship P/U H/D CMI EE

Intake Performed By OS

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bags

Notes: _____

Quality Checks Performed By @

Lab Temp °C 21.5
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 207
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 152 (.139 - .169)
 36mm 175 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 560 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201707D 7/29/19
0.08	SD1013	201707E 7/25/19
0.20	SD3933	201707C 7/24/19
0.08 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By @

Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C ±.2 Serial #: SD3962
 34°C ±.2 Serial #: SD1013
 34°C ±.2 Serial #: SD3933

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By @

Barometric Pressure 1014 Gauge
 ID# 26932 1014 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD3962
0.08	SD1013
0.20	SD3933

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:

OK/OK upon 8/31/2017

JJ Graham 8/31/17
Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC

Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

80-001254
 Stability checks
 8/31/17

INTOXILYZER 8000
 Instrument Initialization
 07:52 08/31/2017

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001256
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:53
Control Test	0.202	08:54
Air Blank	0.000	08:54
Control Test	0.201	08:55
Air Blank	0.000	08:55
Control Test	0.202	08:56
Air Blank	0.000	08:56
Control Test Stats	0.2017	
Average	0.0006	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	

Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001256
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:58
Control Test	0.091	08:59
Air Blank	0.000	08:59
Control Test	0.080	09:00
Air Blank	0.000	09:00
Control Test	0.080	09:01
Air Blank	0.000	09:01
Control Test Stats	0.0803	
Average	0.0006	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001256
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	0.048	09:04
Air Blank	0.000	09:04
Control Test	0.048	09:05
Air Blank	0.000	09:06
Control Test	0.049	09:06
Air Blank	0.000	09:06
Control Test Stats	0.0483	
Average	0.0006	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

DSM

Operator's Signature

DCS

Operator's Signature

8/31/17
 DS