FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange County S/N 80-001258

Date In 5/1/17 Date Out 5/16/17 ☑ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED MAY 16 2017 FDLE Alcohol Testing Program

Intake Performed By DP

- ☐ Registration
- ☑ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☑ Other Static Bag

Notes:

Quality Checks Performed By

- ☑ Lab Temp °C 21.0
- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 179
- ☑ Flow Verification (L/s)

Flow Column # ATP103

32mm 156 (.139 - .169)
36mm 179 (.156 - .190)
53mm 246 (.228 - .278)
103mm 498 (.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 26932

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G3709	296030 / 3/8/18
0.08	DR1279	20161113 / 11/15/18
0.20	DR3856	296040 / 4/5/18
0.08 DGS	N/A	AG626604 / 9/22/18

Flow Calibration Performed By

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column #

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Quality Checks Cont. Performed By DP

Simulator Temperatures °C

External Digital Therm. ID#: 300505

- ☑ 34°C +-.2 Serial #: G3709
- ☑ 34°C +-.2 Serial #: DR1279
- ☑ 34°C +-.2 Serial #: DR3856

Calibration Adjustment Performed By

- ☑ Calibration Adjustment N/A
- ☐ Calibration Adjustment Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

- ☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DP

- ☑ Barometric Pressure ID# 26932
  1015 Gauge
  1008 Instrument

Mouth Alcohol Solution Lot # 2016-C

Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G3709
0.08	DR1279
0.20	DR3856

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Calibration Adjustment
- ☐ Post-Stability Tests
- ☐ Other

Notes/Suggested Service:
QC Runs

JA Graham 5/16/17
Quality Control Review Date

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange County S/N 80-001258

Date In 5/1/17 Date Out _____ ☐Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By BS

- ☐ Registration
- ☒ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other _____

Visual Inspection:

ok Case	ok Handle
ok Dry Gas Holder	ok Feet
ok Keyboard/Plug	ok Back/Plugs
ok Screws tight	ok Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☒ Other Static Bag

Notes: _____

Quality Checks Performed By _____

- ☐ Breath Tube Screen
- ☐ Replace O-Rings
- ☐ Instrument Set Up Verified
- ☐ R-Value _____
- ☐ Flow Verification (L/s)
  - Flow Column # _____
  - 32mm _____ (.139 - .169)
  - 36mm _____ (.156 - .190)
  - 53mm _____ (.228 - .278)
  - 103mm _____ (.447 - .547)
- ☐ Barometric Pressure Check
  - Gauge ID # _____
- ☐ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05		
0.08		
0.20		
0.08 DGS	N/A	

Flow Calibration Performed By _____

RECEIVED
MAY 16 2017
FDLE
Alcohol Testing Program

- ☐ Flow Calibration N/A
- ☐ Flow Calibration Complete
  - Flow Column # _____
  - ☐ 5L/min – 17mm
  - ☐ 15L/min – 53mm
  - ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)
  - Flow Column # _____
  - 32mm _____ (.139 - .169)
  - 36mm _____ (.156 - .190)
  - 53mm _____ (.228 - .278)
  - 103mm _____ (.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Suggested Service

Optical Bench Calibration Performed By _____

- ☐ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By _____

- ☐ Barometric Pressure _____ Gauge
  - ID# _____ _____ Instrument

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.00	
Interferent	
0.05	
0.08	
0.20	

Attachments

- ☐ Form 41
- ☐ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other _____

Notes: QA/QC: Rons

JD 5/16/17

Quality Control Review

- ☐ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☐ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

_____ Date

Revised November 2012

80-001258
Stability Checks
~~5/15/17~~
5/16/17

INTOXILYZER 8000
Instrument Initialization
07:38 05/16/2017

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001258
05/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:27
Control Test	0.080	08:27
Air Blank	0.000	08:27
Control Test	0.079	08:28
Air Blank	0.000	08:28
Control Test	0.079	08:28
Air Blank	0.000	08:29

Control Test Stats

Average	0.0793
Std Dev	0.0006
Rel Std Dev(%)	0.7277

DCS

Operator's Signature

5

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001258
05/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:31
Control Test	0.078	08:32
Air Blank	0.000	08:32
Control Test	0.078	08:33
Air Blank	0.000	08:33
Control Test	0.078	08:34
Air Blank	0.000	08:35

Control Test Stats

Average	0.0780
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001258
05/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.048	08:37
Air Blank	0.000	08:37
Control Test	0.049	08:38
Air Blank	0.000	08:39
Control Test	0.049	08:39
Air Blank	0.000	08:40

Control Test Stats

Average	0.0487
Std Dev	0.0006
Rel Std Dev(%)	1.1863

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001258
05/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:41
Control Test	0.195	08:42
Air Blank	0.000	08:42
Control Test	0.196	08:43
Air Blank	0.000	08:44
Control Test	0.196	08:44
Air Blank	0.000	08:45

Control Test Stats

Average	0.1957
Std Dev	0.0006
Rel Std Dev(%)	0.2951

KDB

Operator's Signature