



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Cocoa PD

S/N 80-001260

Date In 9/14/17

Date Out 9/19/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Lab Temp °C 21.6
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 163
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm .152 (.139 - .169)
 36mm .167 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G2078</u>	<u>201707D</u> <u>7-25-19</u>
0.08	<u>G8149</u>	<u>201707E</u> <u>7-25-19</u>
0.20	<u>G2408</u>	<u>201707C</u> <u>7-24-19</u>
0.08 DGS	<u>N/A</u>	<u>AG626604</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By [Signature]

Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: SD3933

Calibration Adjustment Performed By SP

Calibration Adjustment N/A
 Calibration Adjustment Complete
 Barometric Pressure Gauge 1014 ID # 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	<u>G6621</u>	<u>N/A</u>	<u>N/A</u>
0.040	<u>SD1018</u>	<u>16320</u>	<u>10-21-18</u>
0.100	<u>SD1012</u>	<u>17060</u>	<u>2-14-19</u>
0.200	<u>SD1011</u>	<u>17090</u>	<u>2-24-19</u>
0.300	<u>DR1275</u>	<u>16410</u>	<u>12-19-18</u>
0.080 DGS	<u>N/A</u>	<u>24515089A2</u>	<u>10-5-17</u>

Department Inspection Performed By SP

Barometric Pressure
 ID# 28427 Gauge 1013
1014 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>SD3962</u>
0.08	<u>SD1013</u>
0.20	<u>SD3933</u>

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	<u>SD3962</u>	<u>201707D</u>	<u>7-25-19</u>
0.08	<u>SD1013</u>	<u>201707E</u>	<u>7-25-19</u>
0.20	<u>SD3933</u>	<u>201707C</u>	<u>7-24-19</u>
0.08 DGS	<u>N/A</u>	<u>AG626604</u>	<u>9-22-18</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
QA/OC OK 9/19/2017

[Signature] 9/20/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

STABILITY CHECKS - # 80-001260 COCOA P.D.

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/18/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:09
Control Test	0.049	12:10
Air Blank	0.000	12:10
Control Test	0.049	12:11
Air Blank	0.000	12:12
Control Test	0.049	12:12
Air Blank	0.000	12:13
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

*1/10/16
SP*

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/18/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:14
Control Test	0.081	12:15
Air Blank	0.000	12:15
Control Test	0.082	12:16
Air Blank	0.000	12:16
Control Test	0.082	12:17
Air Blank	0.000	12:18
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

SP
Operator's Signature

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/18/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.200	12:20
Air Blank	0.000	12:20
Control Test	0.202	12:21
Air Blank	0.000	12:22
Control Test	0.202	12:22
Air Blank	0.000	12:23
Control Test Stats		
Average	0.2013	
Std Dev	0.0012	
Rel Std Dev(%)	0.5735	

SP
Operator's Signature

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/18/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.079	12:24
Air Blank	0.000	12:24
Control Test	0.079	12:25
Air Blank	0.000	12:25
Control Test	0.080	12:25
Air Blank	0.000	12:26
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

SPM

SP
Operator's Signature

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000
09/19/2017
SN 80-001260
10:18:12

Auto Calibration
Max Power Res Value = 22
Auto Range Res Value = 11

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12774, Sum Io = 12519

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1010 (-0.0080)
Sample #2 = 0.0690 (0.0480)
Sample #3 = 0.0680 (0.0510)
Sample #4 = 0.0710 (0.0930)
Avg % Abs = 0.0693 (0.0640)
STD DEV = 0.0015 (0.0252)
REL STD DEV = 2.203 (39.312)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1130 (-0.0100)
Sample #2 = 0.1110 (0.0080)
Sample #3 = 0.1040 (0.0150)
Sample #4 = 0.1100 (0.0250)
Avg % Abs = 0.1083 (0.0160)
STD DEV = 0.0038 (0.0085)
REL STD DEV = 3.495 (53.400)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12755, Sum Io = 12510

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8060 (-0.0100)
Sample #2 = 0.7890 (0.0420)
Sample #3 = 0.7950 (0.0430)
Sample #4 = 0.7810 (0.0650)
Avg % Abs = 0.7883 (0.0500)
STD DEV = 0.0070 (0.0130)
REL STD DEV = 0.891 (26.000)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5200 (0.0000)
Sample #2 = 1.5370 (0.0100)
Sample #3 = 1.5170 (0.0190)
Sample #4 = 1.5280 (0.0300)
Avg % Abs = 1.5273 (0.0197)
STD DEV = 0.0100 (0.0100)
REL STD DEV = 0.656 (50.932)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12742, Sum Io = 12507

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9180 (-0.0270)
Sample #2 = 1.8780 (0.0460)
Sample #3 = 1.8770 (0.0320)
Sample #4 = 1.8870 (0.0750)
Avg % Abs = 1.8807 (0.0510)
STD DEV = 0.0055 (0.0219)
REL STD DEV = 0.293 (43.003)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6570 (-0.0020)
Sample #2 = 3.6000 (0.0690)
Sample #3 = 3.6080 (0.0610)
Sample #4 = 3.6120 (0.0810)
Avg % Abs = 3.6067 (0.0703)
STD DEV = 0.0061 (0.0101)
REL STD DEV = 0.169 (14.312)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12732, Sum Io = 12498

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6710 (-0.0090)
Sample #2 = 3.6290 (0.0700)
Sample #3 = 3.6320 (0.0690)
Sample #4 = 3.6590 (0.0650)
Avg % Abs = 3.6400 (0.0680)
STD DEV = 0.0165 (0.0026)
REL STD DEV = 0.454 (3.891)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9100 (-0.0110)
Sample #2 = 6.8940 (0.0670)
Sample #3 = 6.8770 (0.0860)
Sample #4 = 6.9120 (0.0810)
Avg % Abs = 6.8943 (0.0780)
STD DEV = 0.0175 (0.0098)
REL STD DEV = 0.254 (12.627)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12722, Sum Io = 12492

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3300 (-0.0120)
Sample #2 = 5.3070 (0.0420)
Sample #3 = 5.3040 (0.0900)
Sample #4 = 5.2960 (0.1040)
Avg % Abs = 5.3023 (0.0787)
STD DEV = 0.0057 (0.0325)
REL STD DEV = 0.107 (41.335)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.0200 (-0.0140)
Sample #2 = 9.9430 (0.0790)
Sample #3 = 9.9340 (0.1170)
Sample #4 = 9.9530 (0.1090)
Avg % Abs = 9.9433 (0.1017)
STD DEV = 0.0095 (0.0200)
REL STD DEV = 0.096 (19.705)

**** AUTO CAL DATA ****
<<<<< CHANNEL 1 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.069
Std Dev = 0.00 Rel Std Dev = 2.20
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.788
Std Dev = 0.01 Rel Std Dev = 0.89
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.881
Std Dev = 0.01 Rel Std Dev = 0.29
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.640
Std Dev = 0.02 Rel Std Dev = 0.45
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.302
Std Dev = 0.01 Rel Std Dev = 0.11
Zero Order Coef = -152.49
First Order Coef = 2545.24
Second Order Coef = 33.13
Standard Deviation = 24.856771

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.108
Std Dev = 0.00 Rel Std Dev = 3.49
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.527
Std Dev = 0.01 Rel Std Dev = 0.66
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.607
Std Dev = 0.01 Rel Std Dev = 0.17
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.894
Std Dev = 0.02 Rel Std Dev = 0.25
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.943
Std Dev = 0.01 Rel Std Dev = 0.10
Zero Order Coef = -128.02
First Order Coef = 1298.47
Second Order Coef = 15.12
Standard Deviation = 14.908943

<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.000
Std Dev = 0.00
Sol Val = 0.040
% Abs = 0.100
Std Dev = 0.100
Sol Val = 0.200
% Abs = 0.200
Std Dev = 0.200
Sol Val = 0.300
% Abs = 0.300
Std Dev = 0.300

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1
Sample #1 = 3216.00
Sample #2 = 3298.00
Sample #3 = 3184.00
Sample #4 = 3173.00
Average Result = 3218.3333
STD DEV = 69.2122
REL STD DEV = 2.151
**** CHANNEL 2
Sample #1 = 3395.00
Sample #2 = 3418.00
Sample #3 = 3375.00
Sample #4 = 3373.00
Average Result = 3386.6667
STD DEV = 25.4231
REL STD DEV = 0.750
**** CHANNEL 1
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 591
9 um H2O Adjust (mg/l*10,000) = 421
**** AUTO CAL PASS

CAL ADJUSTMENT
#80-001260

CPM

SP

9/20/17
JL

POST CAL ADJUSTMENT STABILITY CHECKS #80-001260

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.050	11:24
Air Blank	0.000	11:24
Control Test	0.049	11:25
Air Blank	0.000	11:26
Control Test	0.050	11:26
Air Blank	0.000	11:27
Control Test Stats	0.0497	
Average	0.0006	
Std Dev	1.1625	
Rel Std Dev(%)		

SP

Operator's Signature

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.081	12:25
Air Blank	0.000	12:26
Control Test	0.080	12:26
Air Blank	0.000	12:27
Control Test	0.081	12:28
Air Blank	0.000	12:28
Control Test Stats	0.0807	
Average	0.0006	
Std Dev	0.7157	
Rel Std Dev(%)		

SP

Operator's Signature

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:13
Control Test	0.201	11:14
Air Blank	0.000	11:14
Control Test	0.202	11:15
Air Blank	0.000	11:16
Control Test	0.200	11:16
Air Blank	0.000	11:17
Control Test Stats	0.2010	
Average	0.0010	
Std Dev	0.4975	
Rel Std Dev(%)		

SP

Operator's Signature

COCOA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001260
09/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:18
Control Test	0.079	11:18
Air Blank	0.000	11:19
Control Test	0.080	11:19
Air Blank	0.000	11:20
Control Test	0.080	11:20
Air Blank	0.000	11:21
Control Test Stats	0.0797	
Average	0.0006	
Std Dev	0.7247	
Rel Std Dev(%)		

DGS

SP

Operator's Signature

SP
1/10/16

SP