


80-001271
 Stability Checks
 10/25/17 

INTOXILYZER 8000
 Instrument Initialization
 06:55 10/25/2017

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001271
 10/25/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:34
Control Test	0.079	07:34
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:35
Control Test	0.080	07:36
Air Blank	0.000	07:36
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

Operator's Signature 

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001271
 10/25/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:39
Control Test	0.204	07:39
Air Blank	0.000	07:40
Control Test	0.202	07:41
Air Blank	0.000	07:41
Control Test	0.203	07:42
Air Blank	0.000	07:42
Control Test Stats		
Average	0.2030	
Std Dev	0.0010	
Rel Std Dev(%)	0.4926	

Operator's Signature 

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001271
 10/25/2017
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	07:43
Control Test	0.080	07:44
Air Blank	0.000	07:45
Control Test	0.079	07:45
Air Blank	0.000	07:46
Control Test	0.079	07:47
Air Blank	0.000	07:47
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

RPM

Operator's Signature 

SEMINOLE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001271
 10/25/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:29
Control Test	0.049	07:30
Air Blank	0.000	07:30
Control Test	0.049	07:31
Air Blank	0.000	07:31
Control Test	0.049	07:32
Air Blank	0.000	07:33
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature 

10/25/17
