FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville S.O. S/N 80-001279

Date In 5/17/17 Date Out 6/1/17 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By

☐ Registration
☒ Annual
☒ Return from CMI
☐ Return from Enforcement Electronics
☐ Other

Visual Inspection:

ok Case
ok Dry Gas Holder
ok Keyboard/Plug
ok Screws tight
ok Handle
ok Feet
ok Back/Plugs
ok Breath Hose

Other Equipment:
☒ Power cord
☐ Printer Cable
☒ Other Static Bag

Notes:

Quality Checks Performed By

☒ Lab Temp °C 20.8
☒ Breath Tube Screen
☐ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value 136
☒ Flow Verification (L/s)

Flow Column # ATP102

32mm 148 (.139 - .169)
36mm 164 (.156 - .190)
53mm 238 (.228 - .278)
103mm 503 (.447 - .547)

☒ Barometric Pressure Check
Gauge ID # 26932
☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	2016030 3/8/18
0.08	SD3964	201611B 11/15/18
0.20	G11621	201604C 4/5/18
0.08 DGS	N/A	AG626604 9/22/18

Flow Calibration Performed By

☒ Flow Calibration N/A
☐ Flow Calibration Complete
Flow Column #
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value
☐ Post Calibration Verification (L/s)

Flow Column #

32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)

RECEIVED
JUN 01 2017
FDLE
Alcohol Testing Program

Maintenance Performed By

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other

Quality Checks Cont. Performed By
Simulator Temperatures °C

External Digital Therm. ID#: 300505
☒ 34°C +-.2 Serial #: DR2035
☒ 34°C +-.2 Serial #: SD3964
☒ 34°C +-.2 Serial #: G11621

Calibration Adjustment Performed By

☒ Calibration Adjustment N/A
☐ Calibration Adjustment Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By

☒ Barometric Pressure 1016 Gauge
ID# 26932 1015 Instrument

Mouth Alcohol Solution Lot # 2016-C
Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	DR2035
0.08	SD3964
0.20	G11621

Attachments

☒ Form 41
☒ Pre-Stability Tests
☐ Flow Calibration
☐ Calibration Adjustment
☐ Post-Stability Tests
☐ Other

Notes/Suggested Service:
Please Change level 2 P.W. to
Something Unique.
QA/QC Ok @ 9am 6/1/2017

Quality Control Review Date 6/1/17

☒ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**

☒ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**

☒ **Conduct an Agency Inspection Before Evidentiary Use**

Revised May 2017
DI-IPS2017R1
FDLE/ATP -048

FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001279

Date In 5/17/17 Date Out ______ ☐Ship ☐P/U ☐H/D ☐CMI ☐EE

RECEIVED
JUN 01 2017
FDLE
Alcohol Testing Program

Intake Performed By ______

☐ Registration
☒ Annual
☒ Return from CMI
☐ Return from Enforcement Electronics
☐ Other ______

Visual Inspection:

ok Case
ok Dry Gas Holder
ok Keyboard/Plug
ok Screws tight
ok Handle
ok Feet
ok Back/Plugs
ok Breath Hose

Other Equipment:

☒ Power cord
☐ Printer Cable
☒ Other Static Bag

Notes: ______

Quality Checks Performed By ______

☐ Lab Temp °C ______
☐ Breath Tube Screen
☐ Replace External O-Rings
☐ Instrument Set Up Verified
☐ R-Value ______
☐ Flow Verification (L/s)

Flow Column # ______

32mm ______ (.139 - .169)
36mm ______ (.156 - .190)
53mm ______ (.228 - .278)
103mm ______ (.447 - .547)

☐ Barometric Pressure Check
Gauge ID # ______
☐ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05		
0.08		
0.20		
0.08 DGS	N/A	

Flow Calibration Performed By ______

☐ Flow Calibration N/A
☐ Flow Calibration Complete
Flow Column # ______
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value ______
☐ Post Calibration Verification (L/s)
Flow Column # ______

32mm ______ (.139 - .169)
36mm ______ (.156 - .190)
53mm ______ (.228 - .278)
103mm ______ (.447 - .547)

Maintenance Performed By ______

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other ______

Quality Checks Cont. Performed By ______

Simulator Temperatures °C

External Digital Therm. ID#: ______
☐ 34°C +-.2 Serial #: ______
☐ 34°C +-.2 Serial #: ______
☐ 34°C +-.2 Serial #: ______

Calibration Adjustment Performed By ______

☐ Calibration Adjustment N/A
☐ Calibration Adjustment Complete
Barometric Pressure Gauge ______ ID # ______

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By ______

☐ Barometric Pressure ID# ______
______ Gauge
______ Instrument

Mouth Alcohol Solution Lot # ______
Acetone Stock Solution Lot # ______

Simulator	Serial Number
0.00	
Interferent	
0.05	
0.08	
0.20	

Attachments

☐ Form 41
☐ Pre-Stability Tests
☐ Flow Calibration
☐ Calibration Adjustment
☐ Post-Stability Tests
☐ Other ______

Notes/Suggested Service: ______

JB

☐ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**

☐ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**

☐ **Conduct an Agency Inspection Before Evidentiary Use**

Quality Control Review Date

80-001279
Stability checks
6/1/17

```
INTOXILYZER 8000
Instrument Initialization
06:21  06/01/2017
```

```
JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000                 SN 80-001279
06/01/2017
Software: 8100.27

Test            g/210L            Time
---------------------------------------
Air Blank       0.000            07:08
Control Test    0.197            07:09
Air Blank       0.000            07:09
Control Test    0.197            07:10
Air Blank       0.000            07:11
Control Test    0.198            07:11
Air Blank       0.000            07:12
Control Test Stats
 Average        0.1973
 Std Dev        0.0006
 Rel Std Dev(%) 0.2926



---------------------------------------
       Operator's Signature
```

```
JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000                 SN 80-001279
06/01/2017
Software: 8100.27

Test            g/210L            Time
---------------------------------------
Air Blank       0.000            07:13
Control Test    0.079            07:14
Air Blank       0.000            07:15
Control Test    0.078            07:15
Air Blank       0.000            07:16
Control Test    0.079            07:17
Air Blank       0.000            07:17
Control Test Stats
 Average        0.0787
 Std Dev        0.0006
 Rel Std Dev(%) 0.7339



---------------------------------------
       Operator's Signature
```

```
JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000                 SN 80-001279
06/01/2017
Software: 8100.27

Test            g/210L            Time
---------------------------------------
Air Blank       0.000            07:18
Control Test    0.048            07:19
Air Blank       0.000            07:19
Control Test    0.048            07:20
Air Blank       0.000            07:21
Control Test    0.049            07:21
Air Blank       0.000            07:22
Control Test Stats
 Average        0.0483
 Std Dev        0.0006
 Rel Std Dev(%) 1.1945



---------------------------------------
       Operator's Signature
```

```
JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000                 SN 80-001279
06/01/2017
Software: 8100.27

Test            g/210L            Time
---------------------------------------
Air Blank       0.000            07:23
Control Test    0.078            07:23
Air Blank       0.000            07:24
Control Test    0.079            07:24
Air Blank       0.000            07:25
Control Test    0.079            07:25
Air Blank       0.000            07:25
Control Test Stats
 Average        0.0787
 Std Dev        0.0006
 Rel Std Dev(%) 0.7339



---------------------------------------
       Operator's Signature
```

DGS