

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001280
 Date In 3/2/17 Date Out 3/16/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 115
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 148 (.139 - .169)
 36mm 467 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR 2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD 3964</u>	<u>201601F</u> <u>1/24/18</u>
0.20	<u>DR 3855</u>	<u>201601C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG-626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1028 Gauge
 ID# 26932 1027 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR 2035</u>
0.08	<u>SD 3964</u>
0.20	<u>DR 3855</u>

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: Please change level 2 P.W. to a unique P.W.
QC V DWS


Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kuitland
 Quality Control Review

3/16/17
 Date

80-001280
 Stability Checks
 3/16/17 

INTOXILYZER 8000
 Instrument Initialization
 07:25 03/16/2017

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001280
 03/16/2017
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	07:54
Control Test	0.047	07:55
Air Blank	0.000	07:56
Control Test	0.048	07:56
Air Blank	0.000	07:57
Control Test	0.048	07:58
Air Blank	0.000	07:58
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel. Std Dev(%)	1.2112	

Operator's Signature 

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001280
 03/16/2017
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:09
Control Test	0.081	08:10
Air Blank	0.000	08:10
Control Test	0.081	08:10
Air Blank	0.000	08:11
Control Test	0.081	08:11
Air Blank	0.000	08:12
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

BSK

Operator's Signature 

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001280
 03/16/2017
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:03
Control Test	0.195	08:04
Air Blank	0.000	08:05
Control Test	0.195	08:05
Air Blank	0.000	08:06
Control Test	0.194	08:06
Air Blank	0.000	08:07
Control Test Stats		
Average	0.1947	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2966	

Operator's Signature 

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001280
 03/16/2017
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	07:51
Control Test	0.080	07:51
Air Blank	0.000	07:52
Control Test	0.081	07:52
Air Blank	0.000	07:53
Control Test	0.081	07:53
Air Blank	0.000	07:54
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7157	

DGS

Operator's Signature 