



INSTRUMENT PROCESSING SHEET

Agency Baker County

S/N 80-001286

Florida Department of Law Enforcement

Date In 10/26/17

DI Completion Date 11/2/17

Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>181</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>164</u> (.139 - .169) 36 mm <u>179</u> (.156 - .190) 53 mm <u>257</u> (.228 - .278) 103 mm <u>529</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G2835</td> <td>201707D 7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 7/25/19</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>201707C 7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AT619605 7/14/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	G2835	201707D 7/25/19	0.080	DR1279	201707E 7/25/19	0.200	SD1025	201707C 7/24/19	0.080 DGS	N/A	AT619605 7/14/18	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.0</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>
Simulator	Serial #	Lot #/Exp															
0.050	G2835	201707D 7/25/19															
0.080	DR1279	201707E 7/25/19															
0.200	SD1025	201707C 7/24/19															
0.080 DGS	N/A	AT619605 7/14/18															
Final Release Date _____ _____																	

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>[Signature]</u> Barometric Pressure ID# <u>28427</u> Gauge <u>709</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2840</td> </tr> <tr> <td>0.050</td> <td>G2835</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G2880	Interferent	G2840	0.050	G2835	0.080	DR1279	0.200	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	G2880																																																												
Interferent	G2840																																																												
0.050	G2835																																																												
0.080	DR1279																																																												
0.200	SD1025																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																													

Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <hr/> Tech Review / Date _____ Admin Review / Date _____
--	---

80-001284
 Stab. l. by Cleets
 11/2/17

BAKER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001286
 11/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:13
Control Test	0.201	06:14
Air Blank	0.000	06:14
Control Test	0.201	06:15
Air Blank	0.000	06:16
Control Test	0.200	06:16
Air Blank	0.000	06:17
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

Operator's Signature

BAKER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001286
 11/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:19
Control Test	0.080	06:20
Air Blank	0.000	06:20
Control Test	0.080	06:21
Air Blank	0.000	06:21
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

BAKER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001286
 11/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:24
Control Test	0.051	06:24
Air Blank	0.000	06:25
Control Test	0.051	06:26
Air Blank	0.000	06:26
Control Test	0.051	06:27
Air Blank	0.000	06:27
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

BAKER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001286
 11/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:29
Control Test	0.081	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:31
Control Test	0.081	06:31
Air Blank	0.000	06:32
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Operator's Signature