INSTRUMENT PROCESSING SHEET

FDLE
Florida Department of Law Enforcement

Agency Baker County S/N 80-001286

Date In 10/26/17 DI Completion Date 11/2/17 ☐Ship ☐P/U ☐H/D ☒CMI ☐EE

Intake Performed By

- ☑ Annual
- ☐ Registration
- ☑ Return from CMI / EE

Visual Inspection:

☑ Case	☑ Handle
☑ Keyboard	☑ Dry Gas Shelf
☑ Feet	☑ Breath Tube
☑ Ports	☑ Screws Tight

Other Equipment/ Accessories:

☑ Power cord	☐ Printer Cable
☑ Static Bag	☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

DEC 14 2017

Alcohol Testing Program

Quality Checks Performed By

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 181
- ☑ Flow Verification (L/s)

Flow Column # ATP105

32 mm	164	(.139 - .169)
36 mm	179	(.156 - .190)
53 mm	257	(.228 - .278)
103 mm	529	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28427

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	G2835	20170TD 7/25/19
0.080	DR1279	20170TE 7/25/19
0.200	SD1025	20170TC 7/24/19
0.080 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By

- ☑ Lab Temp °C 22.0
- External Digital Therm. ID#: 300504
- ☑ 34°C +-.2 Serial #: G2835
- ☑ 34°C +-.2 Serial #: DR1279
- ☑ 34°C +-.2 Serial #: SD1025

Calibration Adjustment Performed By

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By

Barometric Pressure ID# 25427

Gauge 709 Instrument 1018

Mouth Alcohol Solution Lot # 2016-C

Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	G2880
Interferent	G2840
0.050	G2835
0.080	DR1279
0.200	SD1025

Attachments

☑ Form 41	☐ Flow Calibration
☑ Stability Checks	☐ Form 40
☐ Calibration Adjustment	☐ Other _____
☐ Post-Stability Checks	

Notes/Suggested Service: _____

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

12/14/17 Tech Review / Date

12/14/17 Admin Review / Date

80-001286
Stability Checks
11/2/17

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:13
Control Test	0.201	06:14
Air Blank	0.000	06:14
Control Test	0.201	06:15
Air Blank	0.000	06:16
Control Test	0.200	06:16
Air Blank	0.000	06:17

Control Test Stats

Average	0.2007
Std Dev	0.0006
Rel Std Dev(%)	0.2877

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:19
Control Test	0.080	06:20
Air Blank	0.000	06:20
Control Test	0.080	06:21
Air Blank	0.000	06:21

Control Test Stats

Average	0.0803
Std Dev	0.0006
Rel Std Dev(%)	0.7187

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:24
Control Test	0.051	06:24
Air Blank	0.000	06:25
Control Test	0.051	06:26
Air Blank	0.000	06:26
Control Test	0.051	06:27
Air Blank	0.000	06:27

Control Test Stats

Average	0.0510
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:29
Control Test	0.081	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:31
Control Test	0.081	06:31
Air Blank	0.000	06:32

Control Test Stats

Average	0.0810
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DGS

Operator's Signature

12/14/17

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Baker County S/N 80-001286

Date In 10/26/17 DI Completion Date 11/2/17 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By

- ☑ Annual
- ☐ Registration
- ☑ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☑ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes:

Final Release Date

Quality Checks Performed By

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 181
- ☑ Flow Verification (L/s)

Flow Column # ATP105

32 mm	164	(.139 - .169)
36 mm	179	(.156 - .190)
53 mm	257	(.228 - .278)
103 mm	529	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28427

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	G2835	2017070
		7/25/19
0.080	DR1279	2017072
		7/25/19
0.200	SD1025	2017070
		7/24/19
0.080 DGS	N/A	A66196005
		7/14/18

Flow Calibration Performed By

Flow Column #

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Temperature Checks Performed By

- ☑ Lab Temp °C 22.0
- External Digital Therm. ID#: 300504
- ☑ 34°C +-.2 Serial #: G2835
- ☑ 34°C +-.2 Serial #: DR1279
- ☑ 34°C +-.2 Serial #: SD1025

Calibration Adjustment Performed By

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By

Barometric Pressure ID# 28427

Gauge 7019 Instrument 1018

Mouth Alcohol Solution Lot # 2016-C

Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	G2880
Interferent	G2840
0.050	G2835
0.080	DR1279
0.200	SD1025

Attachments

- ☑ Form 41
- ☑ Stability Checks
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other

Notes/Suggested Service:

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

Tech Review / Date

Admin Review / Date

80-001286
Stability Checks
11/2/17

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:13
Control Test	0.201	06:14
Air Blank	0.000	06:14
Control Test	0.201	06:15
Air Blank	0.000	06:16
Control Test	0.200	06:16
Air Blank	0.000	06:17

Control Test Stats

Average	0.2007
Std Dev	0.0006
Rel Std Dev(%)	0.2877

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:19
Control Test	0.080	06:20
Air Blank	0.000	06:20
Control Test	0.080	06:21
Air Blank	0.000	06:21

Control Test Stats

Average	0.0803
Std Dev	0.0006
Rel Std Dev(%)	0.7187

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:24
Control Test	0.051	06:24
Air Blank	0.000	06:25
Control Test	0.051	06:26
Air Blank	0.000	06:26
Control Test	0.051	06:27
Air Blank	0.000	06:27

Control Test Stats

Average	0.0510
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001286
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:29
Control Test	0.081	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:31
Control Test	0.081	06:31
Air Blank	0.000	06:32

Control Test Stats

Average	0.0810
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DGS

Operator's Signature