



80-001288  
 Stability checks  
 11/2/17

INTOXILYZER 8000  
 Instrument Initialization  
 08:30 11/02/2017

UNION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001288  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:37
Control Test	0.079	08:37
Air Blank	0.000	08:38
Control Test	0.079	08:38
Air Blank	0.000	08:39
Control Test	0.080	08:39
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

Operator's Signature

UNION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001288  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.048	08:41
Air Blank	0.000	08:42
Control Test	0.048	08:42
Air Blank	0.000	08:43
Control Test	0.049	08:44
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

*(Signature)*

Operator's Signature

UNION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001288  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.197	08:46
Air Blank	0.000	08:46
Control Test	0.199	08:47
Air Blank	0.000	08:47
Control Test	0.199	08:48
Air Blank	0.000	08:49
Control Test Stats		
Average	0.1983	
Std Dev	0.0012	
Rel Std Dev(%)	0.5822	

*(Signature)*

Operator's Signature

UNION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001288  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.079	08:57
Air Blank	0.000	08:58
Control Test	0.078	08:59
Air Blank	0.000	08:59
Control Test	0.079	09:00
Air Blank	0.000	09:00
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

*(Signature)*

Operator's Signature

11/2/17  
*(Signature)*