



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Bradford County

S/N 80-001290

Date In 1/25/17

Date Out 2/8/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 136
 Flow Verification (L/s)
 Flow Column # ATB102
 32mm 149 (.139 - .169)
 36mm 160 (.156 - .190)
 53mm 237 (.228 - .278)
 103mm 58496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603D 3/8/18
0.08	G8149	201601F 1/26/18
0.20	G11621	201604C 4/5/18
0.08 DGS	N/A	A6626605 9/22/18

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 10 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1010 Gauge
 ID# 28421 1010 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8144
0.20	G11621

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other Form 40

Notes: Volt/Current Diagnostic Fail during inspection. Power fluctuations in building may have accounted for it. Retest no issue. [Signature] in Comp. zone

[Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

2/10/17
Date

80-001290
 Stability Checks
 2/8/17

INTOXILYZER 8000
 Instrument Initialization
 08:17 02/08/2017

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 02/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:58
Control Test	0.080	08:59
Air Blank	0.000	08:59
Control Test	0.079	09:00
Air Blank	0.000	09:01
Control Test	0.081	09:01
Air Blank	0.000	09:02
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

Operator's Signature

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 02/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.050	09:09
Air Blank	0.000	09:10
Control Test	0.050	09:10
Air Blank	0.000	09:11
Control Test	0.050	09:12
Air Blank	0.000	09:12
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BK

Operator's Signature

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 02/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:04
Control Test	0.198	09:04
Air Blank	0.000	09:05
Control Test	0.197	09:06
Air Blank	0.000	09:06
Control Test	0.198	09:07
Air Blank	0.000	09:07
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

Operator's Signature

BRADFORD COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001290
 02/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:50
Control Test	0.080	08:50
Air Blank	0.000	08:51
Control Test	0.079	08:51
Air Blank	0.000	08:52
Control Test	0.078	08:52
Air Blank	0.000	08:52
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

DCS

Operator's Signature

AKS

Operator's Signature