



INSTRUMENT PROCESSING SHEET

Agency Santa Rosa County SOS/N 80-001302

Florida Department of Law Enforcement

Date In 11/9/2017DI Completion Date 11/16/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Annual
 Registration
 Return from CMI / EE

Visual Inspection:
 Case Handle
 Keyboard Dry Gas Shelf
 Feet Breath Tube
 Ports Screws Tight

Other Equipment/ Accessories:
 Power cord Printer Cable
 Static Bag 12V DC Cable

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 177
 Flow Verification (L/s)
 Flow Column # ATP105
 32 mm .160 (.139 - .169)
 36 mm .175 (.156 - .190)
 53 mm .246 (.228 - .278)
 103 mm .500 (.447 - .547)
 Barometric Pressure Check
 Gauge ID # 28427
 Stability Checks

Flow Calibration Performed By _____

Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32 mm _____ (.139 - .169)
 36 mm _____ (.156 - .190)
 53 mm _____ (.228 - .278)
 103 mm _____ (.447 - .547)

Simulator	Serial #	Lot #/Exp
0.050	<u>G2835</u>	201707D 07/25/2019
0.080	<u>DR1279</u>	201707E 07/25/2019
0.200	<u>SD1025</u>	201707C 07/24/2019
0.080 DGS	N/A	<u>AG619605</u> <u>7/14/18</u>

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By [Signature]

Lab Temp °C 21.8
 External Digital Therm. ID#: 300504
 34°C +- .2 Serial #: G2835
 34°C +- .2 Serial #: DR1279
 34°C +- .2 Serial #: SD1025

Final Release Date

NOV 21 2017

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure ID# 28427
 Gauge 1015 Instrument 1007
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	<u>G2880</u>
Interferent	<u>G2840</u>
0.050	<u>G2835</u>
0.080	<u>DR1279</u>
0.200	<u>SD1025</u>

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other _____
 Post-Stability Checks

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

[Signature] 11/20/17 [Signature] 11/21/17
 Tech Review / Date Admin Review / Date

SANTA ROSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001302
 11/16/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:23
Control Test	0.049	08:24
Air Blank	0.000	08:24
Control Test	0.049	08:25
Air Blank	0.000	08:25
Control Test	0.049	08:26
Air Blank	0.000	08:27
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

 Operator's Signature

SANTA ROSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001302
 11/16/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:17
Control Test	0.080	08:18
Air Blank	0.000	08:19
Control Test	0.081	08:19
Air Blank	0.000	08:20
Control Test	0.080	08:20
Air Blank	0.000	08:21
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

P Murphy

 Operator's Signature

SANTA ROSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001302
 11/16/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:13
Control Test	0.200	08:13
Air Blank	0.000	08:14
Control Test	0.199	08:15
Air Blank	0.000	08:15
Control Test	0.198	08:16
Air Blank	0.000	08:16
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

P Murphy

 Operator's Signature

SANTA ROSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001302
 11/16/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:43
Control Test	0.080	08:43
Air Blank	0.000	08:43
Control Test	0.080	08:44
Air Blank	0.000	08:44
Control Test	0.080	08:45
Air Blank	0.000	08:45
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RW

DGS

P Murphy

 Operator's Signature

11/21/17
ee